



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061421\
 Data File : VX022682.D
 Acq On : 14 Jun 2021 09:19
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Title : SW846 8260
 Last Update : Thu Jun 10 12:20:00 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	6931	PASS
75	95	30	60	54.2	17885	PASS
95	95	100	100	100.0	32997	PASS
96	95	5	9	6.3	2070	PASS
173	174	0.00	2	0.8	167	PASS
174	95	50	100	60.8	20064	PASS
175	174	5	9	7.5	1514	PASS
176	174	95	101	95.9	19243	PASS
177	176	5	9	6.9	1320	PASS