

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061422\
 Data File : VX029490.D
 Acq On : 15 Jun 2022 01:02
 Operator : JC/MD
 Sample : N3248-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EXYY8

Quant Time: Jun 15 07:17:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML061322WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 15 07:14:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	343703	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	306528	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	146022	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	108013	40.589	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.180%
7) Chloroethane-d5	1.666	69	85856	51.915	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.820%
11) 1,1-Dichloroethene-d2	2.306	63	163954	34.195	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.400%
21) 2-Butanone-d5	4.458	46	189373	93.857	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.860%
24) Chloroform-d	5.062	84	221088	43.513	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.020%
26) 1,2-Dichloroethane-d4	5.958	65	142637	43.421	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.840%
32) Benzene-d6	5.976	84	462251	48.279	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.560%
36) 1,2-Dichloropropane-d6	7.312	67	136693	48.743	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.480%
41) Toluene-d8	8.653	98	416272	46.116	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.240%
43) trans-1,3-Dichloroprop...	8.952	79	57007	41.225	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.460%
47) 2-Hexanone-d5	9.384	63	143251	94.443	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	196185	47.710	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.420%
66) 1,2-Dichlorobenzene-d4	12.323	152	146900	48.388	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.780%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	18226	7.821	ug/L	# 1
17) trans-1,2-Dichloroethene	3.093	96	8026	3.216	ug/L	91
19) 1,1-Dichloroethane	3.611	63	468887	106.358	ug/L	99
20) cis-1,2-Dichloroethene	4.495	96	632749	224.304	ug/L	95
30) 1,1,1-Trichloroethane	5.385	97	1041602	245.015	ug/L	98
34) Trichloroethene	7.129	95	20888	5.913	ug/L	93
46) Tetrachloroethene	9.275	164	21440	10.573	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061422\
Data File : VX029490.D
Acq On : 15 Jun 2022 01:02
Operator : JC/MD
Sample : N3248-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EXYY8

Quant Time: Jun 15 07:17:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML061322WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 15 07:14:08 2022
Response via : Initial Calibration

