

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061423\
 Data File : VX036155.D
 Acq On : 14 Jun 2023 14:50
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: Jun 15 01:32:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 15 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	404470	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	156935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	33277	5.281	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	10.560%#
35) Dibromofluoromethane	5.391	113	23669	5.687	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	11.380%#
50) Toluene-d8	8.647	98	78106	4.852	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	9.700%#
62) 4-Bromofluorobenzene	11.079	95	26367	6.428	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	12.860%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	10559	4.535	ug/l	94
3) Chloromethane	1.295	50	12167	4.851	ug/l	95
4) Vinyl Chloride	1.374	62	10534	4.460	ug/l	100
5) Bromomethane	1.618	94	6909	4.543	ug/l	86
6) Chloroethane	1.691	64	7495	5.031	ug/l	94
7) Trichlorofluoromethane	1.886	101	14565	4.316	ug/l	100
8) Diethyl Ether	2.136	74	7192	5.114	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	9188	4.480	ug/l	97
10) Methyl Iodide	2.453	142	9382	4.099	ug/l	98
11) Tert butyl alcohol	3.038	59	3792	6.379	ug/l	97
12) 1,1-Dichloroethene	2.319	96	9371	4.596	ug/l	96
13) Acrolein	2.246	56	17921	25.932	ug/l	100
14) Allyl chloride	2.666	41	18258	4.402	ug/l	96
15) Acrylonitrile	3.075	53	34076	24.064	ug/l	99
16) Acetone	2.404	43	33395	24.093	ug/l	99
17) Carbon Disulfide	2.514	76	20875	4.023	ug/l #	94
18) Methyl Acetate	2.709	43	28315	4.860	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	38927	4.812	ug/l	98
20) Methylene Chloride	2.794	84	12753	4.970	ug/l	90
21) trans-1,2-Dichloroethene	3.093	96	10831	4.658	ug/l	95
22) Diisopropyl ether	3.764	45	40028	4.904	ug/l	97
23) Vinyl Acetate	3.727	43	138674	23.240	ug/l	99
24) 1,1-Dichloroethane	3.611	63	20981	4.659	ug/l	98
25) 2-Butanone	4.581	43	47888	23.648	ug/l	99
26) 2,2-Dichloropropane	4.477	77	16106	4.469	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	12919	4.815	ug/l	97
28) Bromochloromethane	4.898	49	12652	5.428	ug/l	96
29) Tetrahydrofuran	5.032	42	30610	23.498	ug/l	98
30) Chloroform	5.093	83	23256	4.932	ug/l	93
31) Cyclohexane	5.471	56	16118	4.478	ug/l	88
32) 1,1,1-Trichloroethane	5.379	97	18157	4.555	ug/l	98
36) 1,1-Dichloropropene	5.696	75	15332	4.406	ug/l	99
37) Ethyl Acetate	4.721	43	18694	4.374	ug/l	99
38) Carbon Tetrachloride	5.678	117	13618	4.054	ug/l	90
39) Methylcyclohexane	7.379	83	15543	4.205	ug/l	94
40) Benzene	6.038	78	46416	4.648	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061423\
 Data File : VX036155.D
 Acq On : 14 Jun 2023 14:50
 Operator : JC/MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Quant Time: Jun 15 01:32:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 15 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	10784	4.525	ug/l	97
42) 1,2-Dichloroethane	6.092	62	20586	4.778	ug/l	100
43) Isopropyl Acetate	6.342	43	29465	4.445	ug/l	99
44) Trichloroethene	7.123	130	11525	4.498	ug/l	96
45) 1,2-Dichloropropane	7.428	63	11935	4.565	ug/l	98
46) Dibromomethane	7.586	93	8731	4.573	ug/l	97
47) Bromodichloromethane	7.824	83	14717	4.134	ug/l	90
48) Methyl methacrylate	7.696	41	15096	4.575	ug/l	99
49) 1,4-Dioxane	7.739	88	6126	94.890	ug/l #	44
51) 4-Methyl-2-Pentanone	8.574	43	93565	23.415	ug/l	99
52) Toluene	8.720	92	28131	4.480	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	14469	3.844	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	16432	4.021	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	12332	4.680	ug/l	98
56) Ethyl methacrylate	9.116	69	18055	4.433	ug/l	95
57) 1,3-Dichloropropane	9.311	76	21150	4.648	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.245	63	48517	22.564	ug/l	99
59) 2-Hexanone	9.433	43	68665	23.164	ug/l	99
60) Dibromochloromethane	9.519	129	9829	3.919	ug/l	99
61) 1,2-Dibromoethane	9.610	107	11883	4.300	ug/l	99
64) Tetrachloroethene	9.275	164	8886	4.473	ug/l	91
65) Chlorobenzene	10.080	112	30740	4.646	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	9942	4.271	ug/l	95
67) Ethyl Benzene	10.195	91	52877	4.524	ug/l	95
68) m/p-Xylenes	10.299	106	40047	8.985	ug/l	97
69) o-Xylene	10.640	106	20151	4.580	ug/l	98
70) Styrene	10.653	104	32573	4.512	ug/l	98
71) Bromoform	10.799	173	6219	6.272	ug/l #	97
73) Isopropylbenzene	10.964	105	49276	4.621	ug/l	100
74) N-amyl acetate	10.842	43	22260	4.589	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.214	83	17765	4.876	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	19316	5.504	ug/l	95
77) Bromobenzene	11.195	156	12404	4.956	ug/l	96
78) n-propylbenzene	11.305	91	56132	4.522	ug/l	99
79) 2-Chlorotoluene	11.360	91	38132	4.851	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	41570	4.619	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	3420	6.702	ug/l #	69
82) 4-Chlorotoluene	11.451	91	43376	4.843	ug/l	97
83) tert-Butylbenzene	11.713	119	38319	4.490	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	41086	4.545	ug/l	99
85) sec-Butylbenzene	11.890	105	45972	4.430	ug/l	99
86) p-Isopropyltoluene	12.006	119	39809	4.532	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	22721	4.759	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	22798	4.628	ug/l	89
89) n-Butylbenzene	12.329	91	31362	4.155	ug/l	98
90) Hexachloroethane	12.536	117	5299	6.101	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	22023	4.710	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	3203	3.840	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	11607	4.367	ug/l	95
94) Hexachlorobutadiene	13.725	225	4123	4.023	ug/l	89
95) Naphthalene	13.774	128	39202	4.179	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	11161	4.230	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061423\
Data File : VX036155.D
Acq On : 14 Jun 2023 14:50
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Quant Time: Jun 15 01:32:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061423W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 15 01:28:15 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.007 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:168 Resp: 214196
Ion Ratio Lower Upper
168 100
99 69.8 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 4.535 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 85 Resp: 10559
Ion Ratio Lower Upper
85 100
87 36.7 16.7 50.1









#3
Chloromethane
Concen: 4.851 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 50 Resp: 12167
Ion Ratio Lower Upper
50 100
52 29.1 25.5 38.3



#4
Vinyl Chloride
Concen: 4.460 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 62 Resp: 10534
Ion Ratio Lower Upper
62 100
64 31.9 25.4 38.0





#5

Bromomethane

Concen: 4.543 ug/l

RT: 1.618 min Scan# 81

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 94 Resp: 6909

Ion Ratio Lower Upper

94 100

96 79.7 74.6 111.8



#6

Chloroethane

Concen: 5.031 ug/l

RT: 1.691 min Scan# 99

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 64 Resp: 7495

Ion Ratio Lower Upper

64 100

66 34.6 25.1 37.7









#7
 Trichlorofluoromethane
 Concen: 4.316 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:101 Resp: 14565
 Ion Ratio Lower Upper
 101 100
 103 64.5 51.4 77.2



#8
 Diethyl Ether
 Concen: 5.114 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50

Tgt Ion: 74 Resp: 7192
 Ion Ratio Lower Upper
 74 100
 45 106.9 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.480 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 9188
Ion Ratio Lower Upper
101 100
85 49.6 37.0 55.6
151 74.1 58.6 87.8



#10
Methyl Iodide
Concen: 4.099 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:142 Resp: 9382
Ion Ratio Lower Upper
142 100
127 53.0 41.6 62.4
141 13.6 11.6 17.4





#11

Tert butyl alcohol

Concen: 6.379 ug/l

RT: 3.038 min Scan# 316

Delta R.T. 0.025 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 59 Resp: 3792

Ion Ratio Lower Upper

59 100

57 8.6 6.0 9.0



#12

1,1-Dichloroethene

Concen: 4.596 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 96 Resp: 9371


Ion Ratio Lower Upper

96 100

61 173.1 144.6 216.8

98 63.8 50.4 75.6





#13

Acrolein

Concen: 25.932 ug/l

RT: 2.246 min Scan# 189

Delta R.T. 0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 56 Resp: 17921

Ion Ratio Lower Upper

56 100

55 71.3 57.1 85.7



#14

Allyl chloride

Concen: 4.402 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 41 Resp: 18258


Ion Ratio Lower Upper

41 100

39 74.4 55.8 83.6

76 29.5 24.3 36.5





#15

Acrylonitrile

Concen: 24.064 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 53 Resp: 34076

Ion Ratio Lower Upper

53 100

52 81.2 66.0 99.0

51 37.6 29.8 44.6



#16

Acetone

Concen: 24.093 ug/l

RT: 2.404 min Scan# 216

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 43 Resp: 33395

Ion Ratio Lower Upper

43 100

58 28.0 21.9 32.9





#17

Carbon Disulfide

Concen: 4.023 ug/l

RT: 2.514 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 76 Resp: 20875

Ion Ratio Lower Upper

76 100

78 11.0 7.0 10.6#



#18

Methyl Acetate

Concen: 4.860 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 43 Resp: 28315

Ion Ratio Lower Upper

43 100

74 21.0 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 4.812 ug/l
RT: 3.117 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 38927
Ion Ratio Lower Upper
73 100
57 22.5 17.3 25.9



#20
Methylene Chloride
Concen: 4.970 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 84 Resp: 12753
Ion Ratio Lower Upper
84 100
49 113.1 103.6 155.4
51 35.0 33.8 50.6
86 62.4 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 4.658 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 10831
Ion Ratio Lower Upper
96 100
61 145.0 119.2 178.8
98 67.7 48.7 73.1



#22
Diisopropyl ether
Concen: 4.904 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 45 Resp: 40028
Ion Ratio Lower Upper
45 100
43 84.4 69.7 104.5
87 24.7 20.8 31.2
59 10.3 9.0 13.4









#23
 Vinyl Acetate
 Concen: 23.240 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.007 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 138674
 Ion Ratio Lower Upper
 43 100
 86 10.3 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 4.659 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50

Tgt Ion: 63 Resp: 20981
 Ion Ratio Lower Upper
 63 100
 98 5.8 3.3 9.8
 100 3.7 2.1 6.5









#25
2-Butanone
Concen: 23.648 ug/l
RT: 4.581 min Scan# 573
Delta R.T. 0.013 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 47888
Ion Ratio Lower Upper
43 100
72 24.0 19.4 29.2



#26
2,2-Dichloropropane
Concen: 4.469 ug/l
RT: 4.477 min Scan# 556
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 77 Resp: 16106
Ion Ratio Lower Upper
77 100
97 21.4 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 4.815 ug/l
 RT: 4.489 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Tgt Ion: 96 Resp: 12919




Ion	Ratio	Lower	Upper
96	100		
61	153.1	0.0	312.2
98	62.1	0.0	129.0




#28
 Bromochloromethane
 Concen: 5.428 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50

Tgt Ion: 49 Resp: 12652

Ion	Ratio	Lower	Upper
49	100		
129	1.1	0.0	3.8
130	68.5	57.5	86.3





#29

Tetrahydrofuran

Concen: 23.498 ug/l

RT: 5.032 min Scan# 645

Delta R.T. 0.013 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 42 Resp: 30610

Ion Ratio Lower Upper

42 100

72 43.7 34.2 51.4

71 39.0 32.2 48.2



#30

Chloroform

Concen: 4.932 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 83 Resp: 23256

Ion Ratio Lower Upper

83 100

85 68.3 50.3 75.5









#31
Cyclohexane
Concen: 4.478 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 56	Resp: 16118
Ion Ratio	Lower Upper
56 100	
69 31.6	23.4 35.2
84 69.5	65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 4.555 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 97	Resp: 18157
Ion Ratio	Lower Upper
97 100	
99 64.8	51.1 76.7
61 48.3	36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 5.281 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. 0.006 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 65 Resp: 33277
 Ion Ratio Lower Upper
 65 100
 67 49.2 0.0 100.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50

Tgt Ion: 114 Resp: 404470
 Ion Ratio Lower Upper
 114 100
 63 22.8 0.0 45.0
 88 18.7 0.0 36.6





#35

Dibromofluoromethane

Concen: 5.687 ug/l

RT: 5.391 min Scan# 705

Delta R.T. 0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:113 Resp: 23669

Ion Ratio Lower Upper

113 100

111 102.5 82.9 124.3

192 16.1 13.3 19.9



#36

1,1-Dichloropropene

Concen: 4.406 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 75 Resp: 15332


Ion Ratio Lower Upper

75 100

110 33.0 17.0 50.9

77 30.8 24.7 37.1





#37

Ethyl Acetate

Concen: 4.374 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 18694

Ion Ratio Lower Upper

43 100

61 14.0 10.6 16.0

70 10.3 8.3 12.5



#38

Carbon Tetrachloride

Concen: 4.054 ug/l

RT: 5.678 min Scan# 753

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 117 Resp: 13618


Ion Ratio Lower Upper

117 100

119 87.2 78.2 117.2

121 26.8 24.4 36.6









#39
Methylcyclohexane
Concen: 4.205 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:	83	Resp:	15543
Ion	Ratio	Lower	Upper
83	100		
55	77.7	64.6	97.0
98	51.8	36.7	55.1



#40
Benzene
Concen: 4.648 ug/l
RT: 6.038 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:	78	Resp:	46416
Ion	Ratio	Lower	Upper
78	100		
77	21.5	19.3	28.9









#41
Methacrylonitrile
Concen: 4.525 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 41	Resp: 10784
Ion Ratio	Lower Upper
41 100	
39 59.4	45.2 67.8
67 64.5	52.6 78.8
52 27.1	22.7 34.1



#42
1,2-Dichloroethane
Concen: 4.778 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 62	Resp: 20586
Ion Ratio	Lower Upper
62 100	
98 8.2	0.0 16.0









#43
Isopropyl Acetate
Concen: 4.445 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:	43	Resp:	29465
Ion	Ratio	Lower	Upper
43	100		
61	18.5	15.2	22.8
87	12.8	9.8	14.6



#44
Trichloroethene
Concen: 4.498 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:	130	Resp:	11525
Ion	Ratio	Lower	Upper
130	100		
95	106.6	0.0	204.4





#45

1,2-Dichloropropane

Concen: 4.565 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 63 Resp: 11935

Ion Ratio Lower Upper

63 100

65 29.5 24.3 36.5



#46

Dibromomethane

Concen: 4.573 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 93 Resp: 8731


Ion Ratio Lower Upper

93 100

95 80.0 66.6 100.0

174 88.2 69.4 104.0









#47
Bromodichloromethane
Concen: 4.134 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 83 Resp: 14717
Ion Ratio Lower Upper
83 100
85 70.5 49.8 74.6
127 9.7 7.1 10.7



#48
Methyl methacrylate
Concen: 4.575 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 41 Resp: 15096
Ion Ratio Lower Upper
41 100
69 72.5 59.2 88.8
39 60.0 47.7 71.5





#49

1,4-Dioxane

Concen: 94.890 ug/l

RT: 7.739 min Scan# 1090

Delta R.T. 0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 88 Resp: 6126

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 27.7 60.1 90.1#



#50

Toluene-d8

Concen: 4.852 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.001 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 98 Resp: 78106

Ion Ratio Lower Upper

98 100

100 66.4 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 23.415 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 43 Resp: 93565

Ion Ratio Lower Upper

43 100

58 37.1 30.4 45.6



#52

Toluene

Concen: 4.480 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 92 Resp: 28131

Ion Ratio Lower Upper

92 100

91 170.7 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 3.844 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 75 Resp: 14469

Ion Ratio Lower Upper

75 100

77 30.6 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 4.021 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 75 Resp: 16432


Ion Ratio Lower Upper

75 100

77 33.5 25.3 37.9

39 58.8 46.7 70.1





#55

1,1,2-Trichloroethane

Concen: 4.680 ug/l

RT: 9.153 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 12332





Ion Ratio Lower Upper

97 100

83 84.6 65.8 98.8

85 53.8 41.6 62.4

99 58.6 47.8 71.8



#56

Ethyl methacrylate

Concen: 4.433 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 69 Resp: 18055


Ion Ratio Lower Upper

69 100

41 73.7 56.8 85.2

39 44.8 32.0 48.0









#57
1,3-Dichloropropane
Concen: 4.648 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 76 Resp: 21150
Ion Ratio Lower Upper
76 100
78 31.0 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 22.564 ug/l
RT: 8.245 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 63 Resp: 48517
Ion Ratio Lower Upper
63 100
106 27.7 22.5 33.7





#59

2-Hexanone

Concen: 23.164 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 43 Resp: 68665

Ion Ratio Lower Upper

43 100

58 51.4 26.2 78.5



#60

Dibromochloromethane

Concen: 3.919 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.001 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50

Tgt Ion: 129 Resp: 9829

Ion Ratio Lower Upper

129 100

127 78.3 38.8 116.4









#61
1,2-Dibromoethane
Concen: 4.300 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:107 Resp: 11883
Ion Ratio Lower Upper
107 100
109 95.3 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 6.428 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 95 Resp: 26367
Ion Ratio Lower Upper
95 100
174 64.6 0.0 135.6
176 64.6 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 362595
Ion Ratio Lower Upper
117 100
82 58.3 46.2 69.4
119 32.0 25.5 38.3



#64
Tetrachloroethene
Concen: 4.473 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:164 Resp: 8886
Ion Ratio Lower Upper
164 100
166 136.9 102.7 154.1
129 111.8 80.4 120.6
131 102.8 76.5 114.7









#65
Chlorobenzene
Concen: 4.646 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 30740
Ion Ratio Lower Upper
112 100
114 31.4 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 4.271 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:131 Resp: 9942
Ion Ratio Lower Upper
131 100
133 98.4 47.1 141.4
119 68.9 32.2 96.6





#67

Ethyl Benzene

Concen: 4.524 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 91 Resp: 52877

Ion Ratio Lower Upper

91 100

106 28.5 24.9 37.3



#68

m/p-Xylenes

Concen: 8.985 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 106 Resp: 40047

Ion Ratio Lower Upper

106 100

91 211.3 165.5 248.3









#69
o-Xylene
Concen: 4.580 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 20151
Ion Ratio Lower Upper
106 100
91 221.6 108.9 326.8



#70
Styrene
Concen: 4.512 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:104 Resp: 32573
Ion Ratio Lower Upper
104 100
78 56.3 43.6 65.4
103 53.3 43.6 65.4









#71
Bromoform
Concen: 6.272 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:173 Resp: 6219
Ion Ratio Lower Upper
173 100
175 46.1 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.006 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion:152 Resp: 156935
Ion Ratio Lower Upper
152 100
115 69.2 43.3 129.9
150 158.2 0.0 345.0





#73

Isopropylbenzene

Concen: 4.621 ug/l

RT: 10.964 min Scan# 1620

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:105 Resp: 49276

Ion Ratio Lower Upper

105 100

120 25.5 12.9 38.6



#74

N-ethyl acetate

Concen: 4.589 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Tgt Ion: 43 Resp: 22260




Ion Ratio Lower Upper


43 100

70 39.3 32.9 49.3

55 25.7 19.3 28.9

61 22.3 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 4.876 ug/l

RT: 11.214 min Scan# 1661

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion: 83 Resp: 17765

Ion Ratio Lower Upper

83 100

131 10.8 5.3 15.8

85 65.0 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 5.504 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 75 Resp: 19316

Ion Ratio Lower Upper

75 100

77 35.1 19.1 57.5





#77

Bromobenzene

Concen: 4.956 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:156 Resp: 12404

Ion Ratio Lower Upper

156 100

77 184.2 88.4 265.3

158 97.5 48.0 144.0



#78

n-propylbenzene

Concen: 4.522 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 91 Resp: 56132

Ion Ratio Lower Upper

91 100

120 22.6 11.5 34.4





#79

2-Chlorotoluene

Concen: 4.851 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 91 Resp: 38132

Ion Ratio Lower Upper

91 100

126 31.8 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 4.619 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 105 Resp: 41570

Ion Ratio Lower Upper

105 100

120 46.7 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 6.702 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 3420
Ion Ratio Lower Upper
75 100
53 159.1 90.5 135.7#
89 48.2 36.9 55.3



#82
4-Chlorotoluene
Concen: 4.843 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion: 91 Resp: 43376
Ion Ratio Lower Upper
91 100
126 26.7 14.2 42.6









#83
tert-Butylbenzene
Concen: 4.490 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion	Ratio	Lower	Upper
119	100		
91	65.9	31.2	93.6
134	24.1	11.7	35.0



#84
1,2,4-Trimethylbenzene
Concen: 4.545 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX036155.D
Acq: 14 Jun 2023 14:50

Tgt Ion	Ratio	Lower	Upper
105	100		
120	43.0	21.2	63.6









#85
 sec-Butylbenzene
 Concen: 4.430 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 45972
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 4.532 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX036155.D
 Acq: 14 Jun 2023 14:50

Tgt Ion:119 Resp: 39809
 Ion Ratio Lower Upper
 119 100
 134 24.5 12.8 38.3
 91 27.6 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 4.759 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 22721

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.1

148 62.8 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 4.628 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion:146 Resp: 22798


Ion Ratio Lower Upper

146 100

111 55.3 21.3 63.7

148 67.6 31.8 95.4





#89

n-Butylbenzene

Concen: 4.155 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 91 Resp: 31362

Ion Ratio Lower Upper

91 100

92 52.6 27.4 82.0

134 24.6 12.6 37.8



#90

Hexachloroethane

Concen: 6.101 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX036155.D




Acq: 14 Jun 2023 14:50


Tgt Ion: 117 Resp: 5299

Ion Ratio Lower Upper

117 100

201 60.6 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 4.710 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 22023

Ion Ratio Lower Upper

146 100

111 47.0 22.4 67.0

148 62.5 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 3.840 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion: 75 Resp: 3203


Ion Ratio Lower Upper

75 100

155 68.9 36.1 108.5

157 96.4 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 4.367 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:180 Resp: 11607

Ion Ratio Lower Upper

180 100

182 90.3 48.3 144.8

145 34.6 18.0 54.0



#94

Hexachlorobutadiene

Concen: 4.023 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50




Tgt Ion:225 Resp: 4123


Ion Ratio Lower Upper

225 100

223 76.5 31.1 93.2

227 68.5 32.5 97.5





#95

Naphthalene

Concen: 4.179 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:128 Resp: 39202

Ion Ratio Lower Upper

128 100

127 13.2 10.3 15.5

129 11.0 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 4.230 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX036155.D

Acq: 14 Jun 2023 14:50

Tgt Ion:180 Resp: 11161

Ion Ratio Lower Upper

180 100

182 95.5 46.9 140.6

145 38.0 17.8 53.5

