

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061423\
 Data File : VX036160.D
 Acq On : 14 Jun 2023 17:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 15 02:13:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 15 02:06:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	272038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	467948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	421008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	204164	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	194849	41.667	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.340%
35) Dibromofluoromethane	5.385	113	139950	44.180	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.360%
50) Toluene-d8	8.647	98	525918	46.383	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.760%
62) 4-Bromofluorobenzene	11.079	95	212683	44.813	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.620%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	1586	0.536	ug/l	97
3) Chloromethane	1.294	50	930	0.292	ug/l	92
4) Vinyl Chloride	1.374	62	1182	0.394	ug/l	99
5) Bromomethane	1.617	94	747	0.387	ug/l	99
6) Chloroethane	1.691	64	499	0.264	ug/l #	83
7) Trichlorofluoromethane	1.892	101	1689	0.394	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.331	101	1545	0.593	ug/l	98
10) Methyl Iodide	2.446	142	963	0.331	ug/l #	83
11) Tert butyl alcohol	2.989	59	177	0.234	ug/l #	78
12) 1,1-Dichloroethene	2.318	96	1097	0.424	ug/l #	73
13) Acrolein	2.239	56	1404	1.600	ug/l	95
14) Allyl chloride	2.666	41	1180	0.224	ug/l #	72
15) Acrylonitrile	3.068	53	3043	1.692	ug/l	97
16) Acetone	2.386	43	2101	1.193	ug/l #	85
17) Carbon Disulfide	2.514	76	12615	1.914	ug/l #	88
18) Methyl Acetate	2.715	43	1716	0.232	ug/l	99
20) Methylene Chloride	2.788	84	1830	0.562	ug/l #	93
21) trans-1,2-Dichloroethene	3.093	96	2501	0.847	ug/l	95
23) Vinyl Acetate	3.727	43	4040	0.533	ug/l #	73
25) 2-Butanone	4.586	43	2010	0.782	ug/l #	51
27) cis-1,2-Dichloroethene	4.489	96	1178	0.346	ug/l	95
28) Bromochloromethane	4.909	49	928	0.313	ug/l #	91
29) Tetrahydrofuran	5.019	42	1103	0.667	ug/l #	79
31) Cyclohexane	5.470	56	1618	0.354	ug/l #	84
36) 1,1-Dichloropropene	5.696	75	3094	0.768	ug/l #	81
38) Carbon Tetrachloride	5.678	117	978	0.252	ug/l #	57
39) Methylcyclohexane	7.385	83	2860	0.669	ug/l	97
40) Benzene	6.043	78	2376	0.206	ug/l	98
42) 1,2-Dichloroethane	6.092	62	1385	0.278	ug/l #	77
44) Trichloroethene	7.129	130	1906	0.643	ug/l	95
46) Dibromomethane	7.586	93	819	0.371	ug/l	93
49) 1,4-Dioxane	7.702	88	84	1.125	ug/l #	51
51) 4-Methyl-2-Pentanone	8.573	43	1934	0.418	ug/l	96
52) Toluene	8.720	92	2619	0.360	ug/l	97
53) t-1,3-Dichloropropene	8.982	75	1582	0.363	ug/l #	84
54) cis-1,3-Dichloropropene	8.366	75	1197	0.253	ug/l #	82

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061423\
 Data File : VX036160.D
 Acq On : 14 Jun 2023 17:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Quant Time: Jun 15 02:13:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 15 02:06:36 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
58) 2-Chloroethyl Vinyl ether	8.250	63	1696	0.682	ug/l	99
59) 2-Hexanone	9.439	43	2472	0.721	ug/l	100
61) 1,2-Dibromoethane	9.610	107	891	0.279	ug/l	96
64) Tetrachloroethene	9.281	164	1910	0.828	ug/l #	73
65) Chlorobenzene	10.079	112	3475	0.452	ug/l	95
67) Ethyl Benzene	10.195	91	5505	0.406	ug/l	97
68) m/p-Xylenes	10.299	106	5329	1.030	ug/l	99
69) o-Xylene	10.640	106	1635	0.320	ug/l	94
70) Styrene	10.658	104	3461	0.413	ug/l	93
71) Bromoform	10.805	173	261	3.520	ug/l #	29
73) Isopropylbenzene	10.963	105	6578	0.474	ug/l	96
74) N-amyl acetate	10.841	43	1534	0.243	ug/l #	83
76) 1,2,3-Trichloropropane	11.079	75	116294	25.470	ug/l #	38
77) Bromobenzene	11.195	156	1437	0.441	ug/l	90
78) n-propylbenzene	11.305	91	13261	0.821	ug/l	99
79) 2-Chlorotoluene	11.360	91	5548	0.543	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	7473	0.638	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.079	75	116294	72.809	ug/l #	5
82) 4-Chlorotoluene	11.457	91	9158	0.786	ug/l	93
83) tert-Butylbenzene	11.713	119	7430	0.669	ug/l	84
84) 1,2,4-Trimethylbenzene	11.750	105	7297	0.620	ug/l	99
85) sec-Butylbenzene	11.890	105	14102	1.045	ug/l	98
86) p-Isopropyltoluene	12.006	119	11347	0.993	ug/l	92
87) 1,3-Dichlorobenzene	11.969	146	5207	0.838	ug/l	94
88) 1,4-Dichlorobenzene	11.969	146	5207	0.812	ug/l	93
89) n-Butylbenzene	12.329	91	15375	1.566	ug/l	98
90) Hexachloroethane	12.536	117	1212	3.464	ug/l	85
91) 1,2-Dichlorobenzene	12.335	146	3929	0.646	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	369	0.340	ug/l	80
93) 1,2,4-Trichlorobenzene	13.585	180	6042	1.747	ug/l	93
94) Hexachlorobutadiene	13.725	225	6009	4.507	ug/l	97
95) Naphthalene	13.774	128	15835	1.298	ug/l	98
96) 1,2,3-Trichlorobenzene	13.963	180	5672	1.653	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Instrument :
MSVOA_X
ClientSampleId :
IBLK









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.001 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 272038
Ion Ratio Lower Upper
168 100
99 68.0 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 0.536 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 85 Resp: 1586
Ion Ratio Lower Upper
85 100
87 31.5 16.7 50.1









#3
Chloromethane
Concen: 0.292 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 50 Resp: 930
Ion Ratio Lower Upper
50 100
52 27.6 25.5 38.3



#4
Vinyl Chloride
Concen: 0.394 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 62 Resp: 1182
Ion Ratio Lower Upper
62 100
64 31.2 25.4 38.0









#5
Bromomethane
Concen: 0.387 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 94 Resp: 747
Ion Ratio Lower Upper
94 100
96 92.1 74.6 111.8



#6
Chloroethane
Concen: 0.264 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 64 Resp: 499
Ion Ratio Lower Upper
64 100
66 22.1 25.1 37.7#









#7
 Trichlorofluoromethane
 Concen: 0.394 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion:101 Resp: 1689
 Ion Ratio Lower Upper
 101 100
 103 61.8 51.4 77.2



#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.593 ug/l
 RT: 2.331 min Scan# 204
 Delta R.T. 0.000 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09

Tgt Ion:101 Resp: 1545
 Ion Ratio Lower Upper
 101 100
 85 42.9 37.0 55.6
 151 72.8 58.6 87.8









#10
Methyl Iodide
Concen: 0.331 ug/l
RT: 2.446 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:142 Resp: 963
Ion Ratio Lower Upper
142 100
127 62.3 41.6 62.4
141 4.5 11.6 17.4#



#11
Tert butyl alcohol
Concen: 0.234 ug/l
RT: 2.989 min Scan# 312
Delta R.T. -0.024 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 59 Resp: 177
Ion Ratio Lower Upper
59 100
57 0.0 6.0 9.0#





#12

1,1-Dichloroethene

Concen: 0.424 ug/l

RT: 2.318 min Scan# 20

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :

IBLK





Tgt Ion: 96 Resp: 1097

Ion Ratio Lower Upper

96 100

61 144.6 144.6 216.8

98 39.9 50.4 75.6#



#13

Acrolein

Concen: 1.600 ug/l

RT: 2.239 min Scan# 189

Delta R.T. 0.000 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion: 56 Resp: 1404

Ion Ratio Lower Upper

56 100

55 75.4 57.1 85.7









#14
 Allyl chloride
 Concen: 0.224 ug/l
 RT: 2.666 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion: 41 Resp: 1180
 Ion Ratio Lower Upper
 41 100
 39 82.7 55.8 83.6
 76 0.0 24.3 36.5#



#15
 Acrylonitrile
 Concen: 1.692 ug/l
 RT: 3.068 min Scan# 325
 Delta R.T. -0.006 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09

Tgt Ion: 53 Resp: 3043
 Ion Ratio Lower Upper
 53 100
 52 82.5 66.0 99.0
 51 32.3 29.8 44.6









#16
Acetone
Concen: 1.193 ug/l
RT: 2.386 min Scan# 216
Delta R.T. -0.018 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 43 Resp: 2101
Ion Ratio Lower Upper
43 100
58 19.4 21.9 32.9#



#17
Carbon Disulfide
Concen: 1.914 ug/l
RT: 2.514 min Scan# 234
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 76 Resp: 12615
Ion Ratio Lower Upper
76 100
78 13.0 7.0 10.6#









#18
Methyl Acetate
Concen: 0.232 ug/l
RT: 2.715 min Scan# 20
Delta R.T. 0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 43 Resp: 1716
Ion Ratio Lower Upper
43 100
74 22.6 17.8 26.8



#20
Methylene Chloride
Concen: 0.562 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 84 Resp: 1830
Ion Ratio Lower Upper
84 100
49 126.1 103.6 155.4
51 29.5 33.8 50.6#
86 56.8 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 0.847 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 96 Resp: 2501
Ion Ratio Lower Upper
96 100
61 157.5 119.2 178.8
98 61.9 48.7 73.1



#23
Vinyl Acetate
Concen: 0.533 ug/l
RT: 3.727 min Scan# 433
Delta R.T. 0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 43 Resp: 4040
Ion Ratio Lower Upper
43 100
86 0.0 7.8 11.8#









#25
2-Butanone
Concen: 0.782 ug/l
RT: 4.586 min Scan# 574
Delta R.T. 0.018 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 43 Resp: 2010
Ion Ratio Lower Upper
43 100
72 0.0 19.4 29.2#



#27
cis-1,2-Dichloroethene
Concen: 0.346 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 96 Resp: 1178
Ion Ratio Lower Upper
96 100
61 157.6 0.0 312.2
98 54.2 0.0 129.0









#28
Bromochloromethane
Concen: 0.313 ug/l
RT: 4.909 min Scan# 61
Delta R.T. 0.012 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 49 Resp: 928
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.8
130 63.9 57.5 86.3



#29
Tetrahydrofuran
Concen: 0.667 ug/l
RT: 5.019 min Scan# 645
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 42 Resp: 1103
Ion Ratio Lower Upper
42 100
72 22.9 34.2 51.4#
71 33.6 32.2 48.2









#31
Cyclohexane
Concen: 0.354 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 56 Resp: 1618
Ion Ratio Lower Upper
56 100
69 40.4 23.4 35.2#
84 95.1 65.9 98.9



#33
1,2-Dichloroethane-d4
Concen: 41.667 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 65 Resp: 194849
Ion Ratio Lower Upper
65 100
67 50.1 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 467948
Ion Ratio Lower Upper
114 100
63 23.1 0.0 45.0
88 17.1 0.0 36.6



#35
Dibromofluoromethane
Concen: 44.180 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion:113 Resp: 139950
Ion Ratio Lower Upper
113 100
111 102.0 82.9 124.3
192 16.7 13.3 19.9









#36
1,1-Dichloropropene
Concen: 0.768 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 75 Resp: 3094
Ion Ratio Lower Upper
75 100
110 16.7 17.0 50.9#
77 27.0 24.7 37.1



#38
Carbon Tetrachloride
Concen: 0.252 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 117 Resp: 978
Ion Ratio Lower Upper
117 100
119 59.2 78.2 117.2#
121 0.0 24.4 36.6#









#39
Methylcyclohexane
Concen: 0.669 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. 0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 83	Resp:	2860
Ion	Ratio	Lower Upper
83	100	
55	76.3	64.6 97.0
98	45.8	36.7 55.1



#40
Benzene
Concen: 0.206 ug/l
RT: 6.043 min Scan# 813
Delta R.T. 0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 78	Resp:	2376
Ion	Ratio	Lower Upper
78	100	
77	25.0	19.3 28.9





#42

1,2-Dichloroethane

Concen: 0.278 ug/l

RT: 6.092 min Scan# 821

Delta R.T. 0.006 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion: 62 Resp: 1385

Ion Ratio Lower Upper

62 100

98 0.0 0.0 16.0



#44

Trichloroethene

Concen: 0.643 ug/l

RT: 7.129 min Scan# 991

Delta R.T. 0.006 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion:130 Resp: 1906

Ion Ratio Lower Upper

130 100

95 107.4 0.0 204.4









#46
Dibromomethane
Concen: 0.371 ug/l
RT: 7.586 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 93 Resp: 819
Ion Ratio Lower Upper
93 100
95 78.0 66.6 100.0
174 78.9 69.4 104.0



#49
1,4-Dioxane
Concen: 1.125 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. -0.030 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 88 Resp: 84
Ion Ratio Lower Upper
88 100
43 101.2 0.0 0.0#
58 116.7 60.1 90.1#





#50

Toluene-d8

Concen: 46.383 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.001 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion: 98 Resp: 525918

Ion Ratio Lower Upper

98 100

100 66.1 53.1 79.7



#51

4-Methyl-2-Pentanone

Concen: 0.418 ug/l

RT: 8.573 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion: 43 Resp: 1934

Ion Ratio Lower Upper

43 100

58 35.5 30.4 45.6





#52

Toluene

Concen: 0.360 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion: 92 Resp: 2619

Ion Ratio Lower Upper

92 100

91 163.8 134.2 201.2



#53

t-1,3-Dichloropropene

Concen: 0.363 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.006 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion: 75 Resp: 1582

Ion Ratio Lower Upper

75 100

77 41.1 25.7 38.5#









#54
 cis-1,3-Dichloropropene
 Concen: 0.253 ug/l
 RT: 8.366 min Scan# 1197
 Delta R.T. 0.000 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion: 75 Resp: 1197
 Ion Ratio Lower Upper
 75 100
 77 41.6 25.3 37.9#
 39 72.3 46.7 70.1#



#58
 2-Chloroethyl Vinyl ether
 Concen: 0.682 ug/l
 RT: 8.250 min Scan# 1175
 Delta R.T. 0.006 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09

Tgt Ion: 63 Resp: 1696
 Ion Ratio Lower Upper
 63 100
 106 27.8 22.5 33.7





#59

2-Hexanone

Concen: 0.721 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.006 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion: 43 Resp: 2472

Ion Ratio Lower Upper

43 100

58 52.4 26.2 78.5



#61

1,2-Dibromoethane

Concen: 0.279 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion: 107 Resp: 891

Ion Ratio Lower Upper

107 100

109 98.3 75.2 112.8









#62
4-Bromofluorobenzene
Concen: 44.813 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 95 Resp: 212683
Ion Ratio Lower Upper
95 100
174 65.9 0.0 135.6
176 62.2 0.0 131.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 117 Resp: 421008
Ion Ratio Lower Upper
117 100
82 58.6 46.2 69.4
119 31.7 25.5 38.3





#64

Tetrachloroethene

Concen: 0.828 ug/l

RT: 9.281 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :

IBLK

Tgt Ion:164 Resp: 1910





Ion Ratio Lower Upper

164 100

166 110.6 102.7 154.1

129 62.3 80.4 120.6#

131 65.5 76.5 114.7#



#65

Chlorobenzene

Concen: 0.452 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion:112 Resp: 3475

Ion Ratio Lower Upper

112 100

114 34.8 25.7 38.5





#67

Ethyl Benzene

Concen: 0.406 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion: 91 Resp: 5505

Ion Ratio Lower Upper

91 100

106 32.7 24.9 37.3



#68

m/p-Xylenes

Concen: 1.030 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion: 106 Resp: 5329

Ion Ratio Lower Upper

106 100

91 205.9 165.5 248.3





#69

o-Xylene

Concen: 0.320 ug/l

RT: 10.640 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion:106 Resp: 1635

Ion Ratio Lower Upper

106 100

91 227.6 108.9 326.8



#70

Styrene

Concen: 0.413 ug/l

RT: 10.658 min Scan# 1570

Delta R.T. 0.006 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09




Tgt Ion:104 Resp: 3461


Ion Ratio Lower Upper

104 100

78 61.8 43.6 65.4

103 51.7 43.6 65.4





#71

Bromoform

Concen: 3.520 ug/l

RT: 10.805 min Scan# 1594

Delta R.T. 0.006 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09





Instrument :

MSVOA_X

ClientSampleId :

IBLK

Tgt Ion:173	Resp:	261
Ion Ratio	Lower	Upper
173	100	
175	0.0	24.1 72.4#
254	0.0	0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Tgt Ion:152	Resp:	204164
Ion Ratio	Lower	Upper
152	100	
115	66.4	43.3 129.9
150	157.8	0.0 345.0









#73
Isopropylbenzene
Concen: 0.474 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:105 Resp: 6578
Ion Ratio Lower Upper
105 100
120 27.6 12.9 38.6



#74
N-ethyl acetate
Concen: 0.243 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 43 Resp: 1534
Ion Ratio Lower Upper
43 100
70 46.2 32.9 49.3
55 25.7 19.3 28.9
61 0.0 18.8 28.2#





#76

1,2,3-Trichloropropane

Concen: 25.470 ug/l

RT: 11.079 min Scan# 1

Delta R.T. -0.158 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion: 75 Resp: 116294

Ion Ratio Lower Upper

75 100

77 1.0 19.1 57.5#



#77

Bromobenzene

Concen: 0.441 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09




Tgt Ion:156 Resp: 1437


Ion Ratio Lower Upper

156 100

77 197.8 88.4 265.3

158 94.8 48.0 144.0









#78
n-propylbenzene
Concen: 0.821 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 91 Resp: 13261
Ion Ratio Lower Upper
91 100
120 23.2 11.5 34.4



#79
2-Chlorotoluene
Concen: 0.543 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 91 Resp: 5548
Ion Ratio Lower Upper
91 100
126 30.1 16.3 48.8









#80
1,3,5-Trimethylbenzene
Concen: 0.638 ug/l
RT: 11.451 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:105 Resp: 7473
Ion Ratio Lower Upper
105 100
120 46.6 24.1 72.3



#81
trans-1,4-Dichloro-2-butene
Concen: 72.809 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.061 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 75 Resp: 116294
Ion Ratio Lower Upper
75 100
53 0.0 90.5 135.7#
89 0.0 36.9 55.3#









#82
4-Chlorotoluene
Concen: 0.786 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 91 Resp: 9158
Ion Ratio Lower Upper
91 100
126 24.5 14.2 42.6



#83
tert-Butylbenzene
Concen: 0.669 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 119 Resp: 7430
Ion Ratio Lower Upper
119 100
91 78.3 31.2 93.6
134 25.7 11.7 35.0





#84

1,2,4-Trimethylbenzene

Concen: 0.620 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion:105 Resp: 7297

Ion Ratio Lower Upper

105 100

120 43.1 21.2 63.6



#85

sec-Butylbenzene

Concen: 1.045 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX036160.D




Acq: 14 Jun 2023 17:09


Tgt Ion:105 Resp: 14102

Ion Ratio Lower Upper

105 100

134 20.5 9.8 29.4





#86

p-Isopropyltoluene

Concen: 0.993 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :

IBLK





Tgt Ion:119 Resp: 11347

Ion Ratio Lower Upper

119 100

134 27.8 12.8 38.3

91 32.2 13.1 39.3



#87

1,3-Dichlorobenzene

Concen: 0.838 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09




Tgt Ion:146 Resp: 5207


Ion Ratio Lower Upper

146 100

111 47.4 22.1 66.1

148 68.8 31.8 95.3









#88
1,4-Dichlorobenzene
Concen: 0.812 ug/l
RT: 11.969 min Scan# 1797
Delta R.T. -0.073 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:146 Resp: 5207
Ion Ratio Lower Upper
146 100
111 47.4 21.3 63.7
148 68.8 31.8 95.4



#89
n-Butylbenzene
Concen: 1.566 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion: 91 Resp: 15375
Ion Ratio Lower Upper
91 100
92 56.0 27.4 82.0
134 24.5 12.6 37.8





#90

Hexachloroethane

Concen: 3.464 ug/l

RT: 12.536 min Scan# 117

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09

Instrument :

MSVOA_X

ClientSampleId :





IBLK

Tgt Ion:117 Resp: 1212

Ion Ratio Lower Upper

117 100

201 53.9 32.9 98.6



#91

1,2-Dichlorobenzene

Concen: 0.646 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX036160.D

Acq: 14 Jun 2023 17:09




Tgt Ion:146 Resp: 3929


Ion Ratio Lower Upper

146 100

111 46.3 22.4 67.0

148 61.2 31.6 95.0









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.340 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. 0.000 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion: 75 Resp: 369
 Ion Ratio Lower Upper
 75 100
 155 53.1 36.1 108.5
 157 72.6 44.9 134.7



#93
 1,2,4-Trichlorobenzene
 Concen: 1.747 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. 0.000 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09

Tgt Ion: 180 Resp: 6042
 Ion Ratio Lower Upper
 180 100
 182 88.8 48.3 144.8
 145 32.7 18.0 54.0









#94
Hexachlorobutadiene
Concen: 4.507 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:	225	Resp:	6009
Ion	Ratio	Lower	Upper
225	100		
223	64.6	31.1	93.2
227	67.2	32.5	97.5



#95
Naphthalene
Concen: 1.298 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX036160.D
Acq: 14 Jun 2023 17:09

Tgt Ion:	128	Resp:	15835
Ion	Ratio	Lower	Upper
128	100		
127	13.7	10.3	15.5
129	10.4	8.8	13.2





#96
 1,2,3-Trichlorobenzene
 Concen: 1.653 ug/l
 RT: 13.963 min Scan# 2112
 Delta R.T. 0.000 min
 Lab File: VX036160.D
 Acq: 14 Jun 2023 17:09

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Tgt Ion:180 Resp: 5672
 Ion Ratio Lower Upper
 180 100
 182 94.0 46.9 140.6
 145 34.9 17.8 53.5

