

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061424\
 Data File : VX041877.D
 Acq On : 14 Jun 2024 14:52
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 15 04:09:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	31743	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	179964	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	160356	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	65673	30.054	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.167%
60) 4-Bromofluorobenzene	11.079	95	67777	27.305	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.033%
63) Toluene-d8	8.647	98	207510	30.247	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.833%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061424\
Data File : VX041877.D
Acq On : 14 Jun 2024 14:52
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 15 04:09:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

