

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061424\
 Data File : VX041886.D
 Acq On : 14 Jun 2024 18:34
 Operator : JC/MD
 Sample : P2893-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-064-IDWGW-20240613

Quant Time: Jun 15 04:15:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	33884	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	180835	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	162629	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	67780	29.058	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.867%
60) 4-Bromofluorobenzene	11.079	95	67667	26.880	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.600%
63) Toluene-d8	8.647	98	208176	29.920	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.733%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	1171	0.570	ug/l	97
15) Acetone	2.392	58	1120	3.124	ug/l	80
37) Trichloroethene	7.129	130	45190	21.042	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061424\
Data File : VX041886.D
Acq On : 14 Jun 2024 18:34
Operator : JC/MD
Sample : P2893-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-064-IDWGW-20240613

Quant Time: Jun 15 04:15:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

