

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061424\
 Data File : VX041878.D
 Acq On : 14 Jun 2024 15:16
 Operator : JC/MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061424

Quant Time: Jun 15 04:10:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	105	0.00
2 M	Dichlorodifluoromethane	1.942	1.907	1.8	101	0.00
3 M	Chloromethane	2.171	2.127	2.0	102	0.00
4 M	Vinyl Chloride	2.123	2.032	4.3	104	0.00
5 M	Bromomethane	1.135	1.048	7.7	101	0.00
6 M	Chloroethane	0.907	1.108	-22.2	128	0.00
7 M	Trichlorofluoromethane	2.674	2.689	-0.6	104	0.00
8 T	Diethyl Ether	1.157	1.080	6.7	101	0.00
9	1,1,2-Trichlorotrifluoroeth	1.818	1.790	1.5	106	0.00
10 M	1,1-Dichloroethene	1.833	1.759	4.0	104	0.00
11	Methyl Iodide	2.386	2.167	9.2	96	0.00
12	Methyl Acetate	2.338	2.238	4.3	101	0.00
13 M	Acrolein	0.471	0.420	10.8	99	0.00
14 M	Acrylonitrile	1.109	1.088	1.9	103	0.00
15 M	Acetone	0.317	0.351	-10.7	114	0.00
16 M	Carbon Disulfide	4.018	3.872	3.6	116	0.00
17	Allyl chloride	2.784	2.827	-1.5	106	0.00
18 M	Methylene Chloride	2.103	1.960	6.8	98	0.00
19 M	trans-1,2-Dichloroethene	1.837	1.762	4.1	104	0.00
20 T	Diisopropyl ether	5.723	5.472	4.4	102	0.00
21 M	1,1-Dichloroethane	3.269	3.042	6.9	101	0.00
22 M	cis-1,2-Dichloroethene	2.223	2.091	5.9	102	0.00
23 M	tert-Butyl Alcohol	0.412	0.479	-16.3	121	0.00
24 M	Methyl tert-Butyl Ether	5.671	5.458	3.8	104	0.00
25 M	Chloroform	3.322	3.171	4.5	102	0.00
26	Cyclohexane	2.840	2.746	3.3	105	0.00
27 s	1,2-Dichloroethane-d4	2.065	1.955	5.3	101	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
29	1,1-Dichloropropene	0.392	0.395	-0.8	108	0.00
30 M	2-Butanone	0.262	0.282	-7.6	107	0.00
31	2,2-Dichloropropane	0.429	0.444	-3.5	110	0.00
32 M	1,1,1-Trichloroethane	0.496	0.488	1.6	104	0.00
33 M	Carbon Tetrachloride	0.428	0.425	0.7	105	0.00
34 M	Benzene	1.304	1.303	0.1	102	0.00
35	Methacrylonitrile	0.263	0.262	0.4	103	0.00
36 M	1,2-Dichloroethane	0.406	0.408	-0.5	103	0.00
37 M	Trichloroethene	0.356	0.354	0.6	106	0.00
38	Methylcyclohexane	0.519	0.530	-2.1	109	0.00
39 M	1,2-Dichloropropane	0.311	0.303	2.6	101	0.00
40	Dibromomethane	0.230	0.229	0.4	104	0.00
41 M	Bromodichloromethane	0.412	0.399	3.2	104	0.00
42 M	Vinyl Acetate	0.876	0.895	-2.2	105	0.00
43	Ethyl Acetate	0.491	0.495	-0.8	106	0.00
44	Isopropyl Acetate	0.735	0.739	-0.5	105	0.00
45 T	1,4-Dioxane	0.009	0.009	0.0	104	0.00
46	Methyl methacrylate	0.359	0.357	0.6	102	0.00
47	n-amyl Acetate	0.659	0.639	3.0	104	0.00
48 M	t-1,3-Dichloropropene	0.424	0.415	2.1	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061424\
 Data File : VX041878.D
 Acq On : 14 Jun 2024 15:16
 Operator : JC/MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061424

Quant Time: Jun 15 04:10:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.476	0.466	2.1	104	0.00
50 M	1,1,2-Trichloroethane	0.334	0.323	3.3	99	0.00
51	Ethyl methacrylate	0.531	0.520	2.1	105	0.00
52	1,3-Dichloropropane	0.546	0.545	0.2	102	0.00
53 M	Dibromochloromethane	0.366	0.351	4.1	104	0.00
54 M	1,2-Dibromoethane	0.352	0.349	0.9	101	0.00
55 M	2-Chloroethyl vinyl ether	0.247	0.232	6.1	93	0.00
56 M	Bromoform	0.298	0.280	6.0	105	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
58 M	4-Methyl-2-Pentanone	0.538	0.553	-2.8	103	0.00
59 M	2-Hexanone	0.415	0.436	-5.1	106	0.00
60 S	4-Bromofluorobenzene	0.464	0.467	-0.6	105	0.00
61 M	Tetrachloroethene	0.377	0.382	-1.3	103	0.00
62 M	Toluene	1.518	1.507	0.7	103	0.00
63 S	Toluene-d8	1.284	1.298	-1.1	103	0.00
64 M	Chlorobenzene	1.034	1.021	1.3	102	0.00
65	1,1,1,2-Tetrachloroethane	0.357	0.351	1.7	104	0.00
66 M	Ethyl Benzene	1.660	1.633	1.6	103	0.00
67 M	m/p-Xylenes	0.668	0.662	0.9	104	0.00
68 M	o-Xylene	0.666	0.654	1.8	102	0.00
69 M	Styrene	1.112	1.107	0.4	103	0.00
70	Isopropylbenzene	1.674	1.668	0.4	103	0.00
71 M	1,1,2,2-Tetrachloroethane	0.579	0.571	1.4	101	0.00
72	1,2,3-Trichloropropane	0.469	0.460	1.9	102	0.00
73	Bromobenzene	0.488	0.477	2.3	103	0.00
74	n-propylbenzene	1.838	1.836	0.1	105	0.00
75	2-Chlorotoluene	1.167	1.153	1.2	102	0.00
76	1,3,5-Trimethylbenzene	1.402	1.390	0.9	102	0.00
77	t-1,4-Dichloro-2-butene	0.174	0.163	6.3	109	0.00
78	4-Chlorotoluene	1.297	1.271	2.0	103	0.00
79	tert-butylbenzene	1.488	1.465	1.5	103	0.00
80	1,2,4-Trimethylbenzene	1.401	1.392	0.6	104	0.00
81	sec-Butylbenzene	1.736	1.725	0.6	104	0.00
82	p-Isopropyltoluene	1.521	1.515	0.4	105	0.00
83 M	1,3-Dichlorobenzene	0.870	0.853	2.0	104	0.00
84 M	1,4-Dichlorobenzene	0.866	0.837	3.3	104	0.00
85	n-Butylbenzene	1.176	1.136	3.4	108	0.00
86 T	Hexachloroethane	0.248	0.238	4.0	107	0.00
87 M	1,2-Dichlorobenzene	0.880	0.853	3.1	102	0.00
88	1,2-Dibromo-3-Chloropropane	0.120	0.113	5.8	104	0.00
89	1,2,4-Trichlorobenzene	0.596	0.559	6.2	107	0.00
90	Hexachlorobutadiene	0.294	0.284	3.4	104	0.00
91 M	Naphthalene	1.800	1.725	4.2	105	0.00
92	1,2,3-Trichlorobenzene	0.617	0.586	5.0	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0