

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061520\
 Data File : VX016764.D
 Acq On : 15 Jun 2020 19:11
 Operator : JC/SP
 Sample : L2993-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AOC50MW002-200610

Quant Time: Jun 16 03:52:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	239642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	375704	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	390501	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203074	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	136182	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
35) Dibromofluoromethane	5.47	113	116292	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
50) Toluene-d8	8.70	98	457285	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
62) 4-Bromofluorobenzene	11.12	95	189524	57.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.94%	
Target Compounds						
16) Acetone	2.42	43	2474	2.023	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061520\
Data File : VX016764.D
Acq On : 15 Jun 2020 19:11
Operator : JC/SP
Sample : L2993-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AOC50MW002-200610

Quant Time: Jun 16 03:52:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

