

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061520\
 Data File : VX016765.D
 Acq On : 15 Jun 2020 19:31
 Operator : JC/SP
 Sample : L2971-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30698

Quant Time: Jun 16 04:25:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	148658	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.84	114	174031	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	103298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	29040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	31548	18.90	ug/l	0.01
Spiked Amount	50.000		Recovery	=	37.80%	
35) Dibromofluoromethane	5.47	113	39309	36.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	73.22%	
50) Toluene-d8	8.70	98	192543	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
62) 4-Bromofluorobenzene	11.12	95	35404	23.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	46.76%	
Target Compounds						
16) Acetone	2.42	43	24576	32.402	ug/l	99
30) Chloroform	5.18	83	8667	3.069	ug/l	98
52) Toluene	8.77	92	12842	4.222	ug/l	97
67) Ethyl Benzene	10.24	91	8428	2.379	ug/l	94
68) m/p-Xylenes	10.34	106	12462	9.276	ug/l	99
69) o-Xylene	10.68	106	2718	2.100	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061520\
Data File : VX016765.D
Acq On : 15 Jun 2020 19:31
Operator : JC/SP
Sample : L2971-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30698

Quant Time: Jun 16 04:25:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

