

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022699.D
 Acq On : 15 Jun 2021 15:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 16 05:25:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72467	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	141641	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57206	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59745	45.84	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.68%		
35) Dibromofluoromethane	5.391	113	44419	49.15	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.30%		
50) Toluene-d8	8.653	98	172474	48.41	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.82%		
62) 4-Bromofluorobenzene	11.085	95	66406	49.29	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.58%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.173	85	42043	49.44	ug/l	97
3) Chloromethane	1.295	50	45047	43.84	ug/l	97
4) Vinyl Chloride	1.374	62	45031	44.60	ug/l	98
5) Bromomethane	1.599	94	30643	55.79	ug/l	99
6) Chloroethane	1.679	64	28019	45.83	ug/l	93
7) Trichlorofluoromethane	1.886	101	70044	48.17	ug/l	97
8) Diethyl Ether	2.136	74	27136	46.20	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.325	101	43906	49.50	ug/l	90
10) Methyl Iodide	2.453	142	40262	40.48	ug/l	94
11) Tert butyl alcohol	2.977	59	57920	191.94	ug/l	98
12) 1,1-Dichloroethene	2.319	96	41649	47.34	ug/l	90
13) Acrolein	2.240	56	29588	233.97	ug/l	99
14) Allyl chloride	2.666	41	82481	48.49	ug/l	94
15) Acrylonitrile	3.069	53	138483	233.66	ug/l	99
16) Acetone	2.386	43	125601	215.22	ug/l	93
17) Carbon Disulfide	2.514	76	94248	47.61	ug/l	99
18) Methyl Acetate	2.709	43	76877	43.76	ug/l	92
19) Methyl tert-butyl Ether	3.123	73	149345	48.92	ug/l	97
20) Methylene Chloride	2.794	84	53966	47.88	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	47014	48.23	ug/l	90
22) Diisopropyl ether	3.764	45	163970	48.71	ug/l #	86
23) Vinyl Acetate	3.727	43	763947	246.26	ug/l #	93
24) 1,1-Dichloroethane	3.611	63	93618	47.93	ug/l	97
25) 2-Butanone	4.562	43	214157	226.27	ug/l	92
26) 2,2-Dichloropropane	4.483	77	77697	52.16	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	57052	48.03	ug/l	95
28) Bromochloromethane	4.904	49	46207	48.92	ug/l #	79
29) Tetrahydrofuran	5.013	42	124615	226.13	ug/l	88
30) Chloroform	5.099	83	93879	48.22	ug/l	96
31) Cyclohexane	5.477	56	72867	46.35	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	77732	47.09	ug/l	96
36) 1,1-Dichloropropene	5.696	75	69216	49.05	ug/l	97
37) Ethyl Acetate	4.721	43	83425	46.51	ug/l #	93
38) Carbon Tetrachloride	5.678	117	63933	52.96	ug/l	97
39) Methylcyclohexane	7.385	83	76679	52.08	ug/l	98
40) Benzene	6.044	78	210238	50.07	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022699.D
 Acq On : 15 Jun 2021 15:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 16 05:25:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	46967	48.86	ug/l	92
42) 1,2-Dichloroethane	6.092	62	77465	50.16	ug/l	96
43) Isopropyl Acetate	6.349	43	141075	49.80	ug/l #	92
44) Trichloroethene	7.129	130	50191	50.92	ug/l	96
45) 1,2-Dichloropropane	7.434	63	56636	49.99	ug/l	97
46) Dibromomethane	7.580	93	37071	48.63	ug/l #	84
47) Bromodichloromethane	7.824	83	71555	48.35	ug/l #	97
48) Methyl methacrylate	7.696	41	68118	50.09	ug/l	88
49) 1,4-Dioxane	7.665	88	21935	847.48	ug/l #	93
51) 4-Methyl-2-Pentanone	8.580	43	443903	246.87	ug/l	90
52) Toluene	8.720	92	131148	50.19	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	81866	46.69	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	89384	47.50	ug/l	93
55) 1,1,2-Trichloroethane	9.153	97	53861	50.37	ug/l	96
56) Ethyl methacrylate	9.116	69	89502	46.45	ug/l #	84
57) 1,3-Dichloropropane	9.311	76	97144	52.24	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.245	63	239465	248.43	ug/l	95
59) 2-Hexanone	9.433	43	343064	249.71	ug/l	89
60) Dibromochloromethane	9.525	129	49336	47.17	ug/l	99
61) 1,2-Dibromoethane	9.610	107	54719	51.27	ug/l	100
64) Tetrachloroethene	9.275	164	43616	52.95	ug/l #	89
65) Chlorobenzene	10.080	112	137370	50.60	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	47598	48.49	ug/l	98
67) Ethyl Benzene	10.195	91	257578	51.54	ug/l	99
68) m/p-Xylenes	10.305	106	190051	104.98	ug/l	95
69) o-Xylene	10.647	106	93307	53.19	ug/l	97
70) Styrene	10.659	104	158671	49.01	ug/l	96
71) Bromoform	10.799	173	29578	46.28	ug/l #	99
73) Isopropylbenzene	10.964	105	250428	53.31	ug/l	99
74) N-amyl acetate	10.848	43	128381	50.82	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.214	83	86231	51.61	ug/l	98
76) 1,2,3-Trichloropropane	11.244	75	99953	65.90	ug/l	88
77) Bromobenzene	11.201	156	51154	50.98	ug/l	83
78) n-propylbenzene	11.305	91	304439	53.66	ug/l	97
79) 2-Chlorotoluene	11.366	91	176367	51.37	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	209819	53.62	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.025	75	27114	46.72	ug/l	93
82) 4-Chlorotoluene	11.457	91	205977	52.43	ug/l	96
83) tert-Butylbenzene	11.720	119	194541	52.94	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	213634	53.83	ug/l	99
85) sec-Butylbenzene	11.896	105	260916	54.36	ug/l	98
86) p-Isopropyltoluene	12.012	119	216464	54.63	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	99877	52.44	ug/l	96
88) 1,4-Dichlorobenzene	12.043	146	101487	51.76	ug/l	96
89) n-Butylbenzene	12.335	91	216596	57.09	ug/l	96
90) Hexachloroethane	12.543	117	32419	49.28	ug/l	72
91) 1,2-Dichlorobenzene	12.335	146	98213	52.92	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	17973	45.95	ug/l	75
93) 1,2,4-Trichlorobenzene	13.591	180	59666	53.59	ug/l	97
94) Hexachlorobutadiene	13.725	225	20800	52.10	ug/l	96
95) Naphthalene	13.780	128	235989	48.58	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	59145	54.49	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
Data File : VX022699.D
Acq On : 15 Jun 2021 15:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Jun 16 05:25:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050





Tgt Ion:168 Resp: 72467
Ion Ratio Lower Upper
168 100
99 69.3 45.8 68.6#




#2
Dichlorodifluoromethane
Concen: 49.44 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.007 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01



Tgt Ion: 85 Resp: 42043
Ion Ratio Lower Upper
85 100
87 31.1 16.4 49.2









#3
Chloromethane
Concen: 43.84 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 45047
Ion Ratio Lower Upper
50 100
52 31.5 26.4 39.6



#4
Vinyl Chloride
Concen: 44.60 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 62 Resp: 45031
Ion Ratio Lower Upper
62 100
64 30.8 25.7 38.5









#5
Bromomethane
Concen: 55.79 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 30643
Ion Ratio Lower Upper
94 100
96 94.5 76.4 114.6



#6
Chloroethane
Concen: 45.83 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 64 Resp: 28019
Ion Ratio Lower Upper
64 100
66 35.5 25.4 38.0









#7
Trichlorofluoromethane
Concen: 48.17 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 70044
Ion Ratio Lower Upper
101 100
103 63.9 53.2 79.8



#8
Diethyl Ether
Concen: 46.20 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 74 Resp: 27136
Ion Ratio Lower Upper
74 100
45 104.6 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.50 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 43906
Ion Ratio Lower Upper
101 100
85 48.0 34.9 52.3
151 66.4 61.4 92.2



#10
Methyl Iodide
Concen: 40.48 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:142 Resp: 40262
Ion Ratio Lower Upper
142 100
127 45.9 33.3 49.9
141 13.5 11.4 17.0









#11
Tert butyl alcohol
Concen: 191.94 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 57920
Ion Ratio Lower Upper
59 100
57 10.5 7.8 11.8



#12
1,1-Dichloroethene
Concen: 47.34 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 96 Resp: 41649
Ion Ratio Lower Upper
96 100
61 172.1 124.6 187.0
98 64.9 50.6 76.0









#13
Acrolein
Concen: 233.97 ug/l
RT: 2.240 min Scan# 189
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 29588
Ion Ratio Lower Upper
56 100
55 71.7 56.4 84.6



#14
Allyl chloride
Concen: 48.49 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 41 Resp: 82481
Ion Ratio Lower Upper
41 100
39 67.5 51.6 77.4
76 32.2 30.5 45.7









#15
Acrylonitrile
Concen: 233.66 ug/l
RT: 3.069 min Scan# 325
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	53	Resp:	138483
Ion	Ratio	Lower	Upper
53	100		
52	81.1	65.3	97.9
51	35.5	28.2	42.2



#16
Acetone
Concen: 215.22 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:	43	Resp:	125601
Ion	Ratio	Lower	Upper
43	100		
58	30.6	27.8	41.8









#17
Carbon Disulfide
Concen: 47.61 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 94248
Ion Ratio Lower Upper
76 100
78 9.3 7.2 10.8



#18
Methyl Acetate
Concen: 43.76 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 43 Resp: 76877
Ion Ratio Lower Upper
43 100
74 23.2 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 48.92 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 149345
Ion Ratio Lower Upper
73 100
57 22.6 17.1 25.7



#20
Methylene Chloride
Concen: 47.88 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 84 Resp: 53966
Ion Ratio Lower Upper
84 100
49 123.3 88.6 132.8
51 37.4 27.0 40.6
86 62.7 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 48.23 ug/l
RT: 3.093 min Scan# 329
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 96 Resp: 47014




Ion	Ratio	Lower	Upper
96	100		
61	152.2	107.7	161.5
98	63.5	50.4	75.6




#22
Diisopropyl ether
Concen: 48.71 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 45 Resp: 163970

Ion	Ratio	Lower	Upper
45	100		
43	87.0	58.0	87.0#
87	25.9	25.4	38.0
59	10.9	10.2	15.2









#23
 Vinyl Acetate
 Concen: 246.26 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. 0.000 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 763947
 Ion Ratio Lower Upper
 43 100
 86 9.8 9.9 14.9#



#24
 1,1-Dichloroethane
 Concen: 47.93 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01

Tgt Ion: 63 Resp: 93618
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.8 11.2
 100 3.8 2.5 7.5









#25
2-Butanone
Concen: 226.27 ug/l
RT: 4.562 min Scan# 570
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 214157
Ion Ratio Lower Upper
43 100
72 25.4 23.7 35.5



#26
2,2-Dichloropropane
Concen: 52.16 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 77 Resp: 77697
Ion Ratio Lower Upper
77 100
97 22.2 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 48.03 ug/l

RT: 4.495 min Scan# 559





Delta R.T. 0.000 min

Lab File: VX022699.D

Acq: 15 Jun 2021 15:01

Tgt Ion: 96 Resp: 57052

Ion	Ratio	Lower	Upper
96	100		
61	150.6	0.0	285.6
98	62.6	0.0	129.6



#28

Bromochloromethane

Concen: 48.92 ug/l

RT: 4.904 min Scan# 626




Delta R.T. -0.000 min


Lab File: VX022699.D

Acq: 15 Jun 2021 15:01

Tgt Ion: 49 Resp: 46207

Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	4.0
130	65.4	68.2	102.2#





#29

Tetrahydrofuran

Concen: 226.13 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX022699.D

Acq: 15 Jun 2021 15:01





Tgt Ion: 42 Resp: 124615

Ion Ratio Lower Upper

42 100

72 43.0 41.7 62.5

71 40.1 38.4 57.6



#30

Chloroform

Concen: 48.22 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX022699.D




Acq: 15 Jun 2021 15:01


Tgt Ion: 83 Resp: 93879

Ion Ratio Lower Upper

83 100

85 62.0 52.0 78.0









#31
Cyclohexane
Concen: 46.35 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	56	Resp:	72867
Ion	Ratio	Lower	Upper
56	100		
69	28.1	25.3	37.9
84	83.8	71.4	107.0



#32
1,1,1-Trichloroethane
Concen: 47.09 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:	97	Resp:	77732
Ion	Ratio	Lower	Upper
97	100		
99	64.1	51.5	77.3
61	48.2	34.2	51.4









#33
 1,2-Dichloroethane-d4
 Concen: 45.84 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. -0.006 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 65 Resp: 59745
 Ion Ratio Lower Upper
 65 100
 67 52.0 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. 0.000 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01

Tgt Ion: 114 Resp: 141641
 Ion Ratio Lower Upper
 114 100
 63 22.4 0.0 40.8
 88 16.0 0.0 32.8









#35
Dibromofluoromethane
Concen: 49.15 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 44419
Ion Ratio Lower Upper
113 100
111 102.9 81.9 122.9
192 15.0 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 49.05 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 75 Resp: 69216
Ion Ratio Lower Upper
75 100
110 33.5 18.1 54.4
77 31.2 25.3 37.9





#37

Ethyl Acetate

Concen: 46.51 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.006 min

Lab File: VX022699.D

Acq: 15 Jun 2021 15:01





Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43	Resp: 83425
Ion Ratio	Lower Upper
43 100	
61 13.2	12.8 19.2
70 10.2	10.5 15.7#



#38

Carbon Tetrachloride

Concen: 52.96 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.000 min

Lab File: VX022699.D

Acq: 15 Jun 2021 15:01

Tgt Ion: 117	Resp: 63933
Ion Ratio	Lower Upper
117 100	
119 91.7	76.2 114.2
121 29.7	24.8 37.2









#39
Methylcyclohexane
Concen: 52.08 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 76679
Ion Ratio Lower Upper
83 100
55 75.7 58.9 88.3
98 48.1 38.1 57.1



#40
Benzene
Concen: 50.07 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 78 Resp: 210238
Ion Ratio Lower Upper
78 100
77 22.8 18.9 28.3









#41
Methacrylonitrile
Concen: 48.86 ug/l
RT: 4.928 min Scan# 630
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	46967
Ion	Ratio	Lower	Upper
41	100		
39	54.5	42.5	63.7
67	68.8	65.0	97.6
52	28.9	25.3	37.9



#42
1,2-Dichloroethane
Concen: 50.16 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:	62	Resp:	77465
Ion	Ratio	Lower	Upper
62	100		
98	8.8	0.0	20.2









#43
Isopropyl Acetate
Concen: 49.80 ug/l
RT: 6.349 min Scan# 863
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 141075
Ion Ratio Lower Upper
43 100
61 19.0 18.2 27.4
87 11.8 12.5 18.7#



#44
Trichloroethene
Concen: 50.92 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:130 Resp: 50191
Ion Ratio Lower Upper
130 100
95 103.1 0.0 198.0









#45
1,2-Dichloropropane
Concen: 49.99 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 63 Resp: 56636
Ion Ratio Lower Upper
63 100
65 30.2 25.7 38.5



#46
Dibromomethane
Concen: 48.63 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 93 Resp: 37071
Ion Ratio Lower Upper
93 100
95 82.5 66.8 100.2
174 75.8 83.1 124.7#









#47
Bromodichloromethane
Concen: 48.35 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 71555
Ion Ratio Lower Upper
83 100
85 62.5 51.4 77.0
127 6.4 6.8 10.2#



#48
Methyl methacrylate
Concen: 50.09 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 41 Resp: 68118
Ion Ratio Lower Upper
41 100
69 79.9 76.0 114.0
39 60.3 44.6 67.0









#49
1,4-Dioxane
Concen: 847.48 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 88 Resp: 21935
Ion Ratio Lower Upper
88 100
43 37.4 23.2 34.8#
58 71.8 55.0 82.4



#50
Toluene-d8
Concen: 48.41 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 98 Resp: 172474
Ion Ratio Lower Upper
98 100
100 65.4 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 246.87 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 443903
Ion Ratio Lower Upper
43 100
58 37.7 35.4 53.0



#52
Toluene
Concen: 50.19 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 92 Resp: 131148
Ion Ratio Lower Upper
92 100
91 169.8 137.1 205.7









#53
 t-1,3-Dichloropropene
 Concen: 46.69 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. 0.000 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 75 Resp: 81866
 Ion Ratio Lower Upper
 75 100
 77 31.5 24.9 37.3



#54
 cis-1,3-Dichloropropene
 Concen: 47.50 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.006 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01

Tgt Ion: 75 Resp: 89384
 Ion Ratio Lower Upper
 75 100
 77 33.5 25.7 38.5
 39 50.8 35.1 52.7





#55

1,1,2-Trichloroethane

Concen: 50.37 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.006 min

Lab File: VX022699.D

Acq: 15 Jun 2021 15:01

Tgt Ion: 97 Resp: 53861





Ion Ratio Lower Upper

97 100

83 91.8 69.3 103.9

85 60.0 44.3 66.5

99 62.4 50.3 75.5



#56

Ethyl methacrylate

Concen: 46.45 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VX022699.D

Acq: 15 Jun 2021 15:01




Tgt Ion: 69 Resp: 89502


Ion Ratio Lower Upper

69 100

41 68.3 46.8 70.2

39 42.0 24.7 37.1









#57

1,3-Dichloropropane
Concen: 52.24 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 97144
Ion Ratio Lower Upper
76 100
78 31.5 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 248.43 ug/l
RT: 8.245 min Scan# 1174
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 63 Resp: 239465
Ion Ratio Lower Upper
63 100
106 26.5 23.5 35.3









#59
2-Hexanone
Concen: 249.71 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 343064
Ion Ratio Lower Upper
43 100
58 53.3 30.9 92.8



#60
Dibromochloromethane
Concen: 47.17 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:129 Resp: 49336
Ion Ratio Lower Upper
129 100
127 77.5 38.4 115.2









#61
1,2-Dibromoethane
Concen: 51.27 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 54719
Ion Ratio Lower Upper
107 100
109 94.7 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 49.29 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 95 Resp: 66406
Ion Ratio Lower Upper
95 100
174 62.9 0.0 154.6
176 60.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 130265
Ion Ratio Lower Upper
117 100
82 58.1 47.4 71.2
119 31.7 26.6 39.8



#64
Tetrachloroethene
Concen: 52.95 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:164 Resp: 43616
Ion Ratio Lower Upper
164 100
166 126.4 103.4 155.2
129 107.6 70.8 106.2#
131 104.4 72.5 108.7









#65
Chlorobenzene
Concen: 50.60 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 137370
Ion Ratio Lower Upper
112 100
114 33.2 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 48.49 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:131 Resp: 47598
Ion Ratio Lower Upper
131 100
133 95.7 47.7 143.1
119 67.9 32.1 96.5









#67
Ethyl Benzene
Concen: 51.54 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 257578
Ion Ratio Lower Upper
91 100
106 30.3 24.6 37.0



#68
m/p-Xylenes
Concen: 104.98 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:106 Resp: 190051
Ion Ratio Lower Upper
106 100
91 210.8 162.8 244.2









#69
o-Xylene
Concen: 53.19 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 93307
Ion Ratio Lower Upper
106 100
91 221.2 108.1 324.2



#70
Styrene
Concen: 49.01 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:104 Resp: 158671
Ion Ratio Lower Upper
104 100
78 54.3 40.0 60.0
103 53.5 43.4 65.0









#71
Bromoform
Concen: 46.28 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 29578
Ion Ratio Lower Upper
173 100
175 48.9 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:152 Resp: 57206
Ion Ratio Lower Upper
152 100
115 98.0 41.1 123.3
150 177.7 0.0 349.0









#73
Isopropylbenzene
Concen: 53.31 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 250428
Ion Ratio Lower Upper
105 100
120 25.7 13.2 39.6



#74
N-ethyl acetate
Concen: 50.82 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 43 Resp: 128381
Ion Ratio Lower Upper
43 100
70 40.7 39.0 58.6
55 23.0 21.5 32.3
61 23.5 22.1 33.1









#75

1,1,2,2-Tetrachloroethane
Concen: 51.61 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 86231
Ion Ratio Lower Upper
83 100
131 8.4 5.1 15.2
85 64.5 31.8 95.4




#76

1,2,3-Trichloropropane
Concen: 65.90 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 75 Resp: 99953
Ion Ratio Lower Upper
75 100
77 31.1 19.1 57.5









#77
Bromobenzene
Concen: 50.98 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 51154
Ion Ratio Lower Upper
156 100
77 201.7 82.8 248.6
158 97.7 49.1 147.4



#78
n-propylbenzene
Concen: 53.66 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 91 Resp: 304439
Ion Ratio Lower Upper
91 100
120 21.6 11.5 34.5









#79
2-Chlorotoluene
Concen: 51.37 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 176367
Ion Ratio Lower Upper
91 100
126 31.6 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 53.62 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 105 Resp: 209819
Ion Ratio Lower Upper
105 100
120 46.9 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 46.72 ug/l
RT: 11.025 min Scan# 1630
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	75	Resp:	27114
Ion	Ratio	Lower	Upper
75	100		
53	103.4	74.5	111.7
89	46.8	37.9	56.9



#82
4-Chlorotoluene
Concen: 52.43 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:	91	Resp:	205977
Ion	Ratio	Lower	Upper
91	100		
126	27.5	14.9	44.5









#83
tert-Butylbenzene
Concen: 52.94 ug/l
RT: 11.720 min Scan# 1744
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 194541
Ion Ratio Lower Upper
119 100
91 65.0 27.9 83.7
134 22.8 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 53.83 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:105 Resp: 213634
Ion Ratio Lower Upper
105 100
120 44.2 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 54.36 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 260916
 Ion Ratio Lower Upper
 105 100
 134 18.8 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 54.63 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022699.D
 Acq: 15 Jun 2021 15:01

Tgt Ion:119 Resp: 216464
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.9 38.7
 91 26.6 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 52.44 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 99877
Ion Ratio Lower Upper
146 100
111 46.3 20.2 60.5
148 64.1 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 51.76 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:146 Resp: 101487
Ion Ratio Lower Upper
146 100
111 44.6 19.6 58.8
148 62.1 31.6 95.0









#89
n-Butylbenzene
Concen: 57.09 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 216596
Ion Ratio Lower Upper
91 100
92 52.5 27.4 82.0
134 22.8 13.2 39.5



#90
Hexachloroethane
Concen: 49.28 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. 0.001 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 117 Resp: 32419
Ion Ratio Lower Upper
117 100
201 56.1 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 52.92 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 98213
Ion Ratio Lower Upper
146 100
111 46.6 20.8 62.2
148 63.1 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 45.95 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion: 75 Resp: 17973
Ion Ratio Lower Upper
75 100
155 69.7 45.3 135.9
157 86.3 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 53.59 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 59666
Ion Ratio Lower Upper
180 100
182 95.2 47.7 143.1
145 35.7 15.0 45.0



#94
Hexachlorobutadiene
Concen: 52.10 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.006 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:225 Resp: 20800
Ion Ratio Lower Upper
225 100
223 64.4 31.1 93.2
227 67.6 31.6 95.0









#95
Naphthalene
Concen: 48.58 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128	Resp: 235989
Ion Ratio	Lower Upper
128	100
127	12.5 10.3 15.5
129	11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 54.49 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX022699.D
Acq: 15 Jun 2021 15:01

Tgt Ion:180	Resp: 59145
Ion Ratio	Lower Upper
180	100
182	94.5 47.6 142.9
145	37.2 16.1 48.3

