

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022702.D
 Acq On : 15 Jun 2021 16:30
 Operator : JC/MD
 Sample : VX0615WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0615WBS01

Quant Time: Jun 16 05:26:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	71762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137264	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54349	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58782	45.55	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.10%		
35) Dibromofluoromethane	5.391	113	42848	48.93	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.86%		
50) Toluene-d8	8.653	98	171772	49.75	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.50%		
62) 4-Bromofluorobenzene	11.085	95	63463	48.61	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.22%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.173	85	15419	18.31	ug/l	97
3) Chloromethane	1.294	50	16706	16.42	ug/l	99
4) Vinyl Chloride	1.374	62	16690	16.69	ug/l	99
5) Bromomethane	1.599	94	9093	16.72	ug/l	98
6) Chloroethane	1.685	64	10205	16.86	ug/l	89
7) Trichlorofluoromethane	1.886	101	25405	17.64	ug/l	96
8) Diethyl Ether	2.136	74	10323	17.75	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.331	101	15866	18.06	ug/l	88
10) Methyl Iodide	2.453	142	13012	13.21	ug/l #	90
11) Tert butyl alcohol	2.977	59	24608	82.35	ug/l	99
12) 1,1-Dichloroethene	2.319	96	15673	17.99	ug/l	96
13) Acrolein	2.239	56	12700	101.41	ug/l	100
14) Allyl chloride	2.666	41	29075	17.26	ug/l	92
15) Acrylonitrile	3.068	53	52789	89.95	ug/l	100
16) Acetone	2.386	43	48166	83.35	ug/l	97
17) Carbon Disulfide	2.514	76	31252	15.94	ug/l	97
18) Methyl Acetate	2.709	43	29749	17.10	ug/l	93
19) Methyl tert-butyl Ether	3.123	73	54028	17.87	ug/l	99
20) Methylene Chloride	2.794	84	20379	18.09	ug/l	87
21) trans-1,2-Dichloroethene	3.093	96	16299	16.88	ug/l	89
22) Diisopropyl ether	3.770	45	59252	17.78	ug/l	96
23) Vinyl Acetate	3.727	43	274062	89.21	ug/l	95
24) 1,1-Dichloroethane	3.611	63	34415	17.79	ug/l	97
25) 2-Butanone	4.568	43	81097	86.53	ug/l #	89
26) 2,2-Dichloropropane	4.483	77	27930	18.93	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	20731	17.62	ug/l	97
28) Bromochloromethane	4.910	49	17375	18.58	ug/l #	74
29) Tetrahydrofuran	5.019	42	47899	87.77	ug/l	88
30) Chloroform	5.105	83	34559	17.92	ug/l	95
31) Cyclohexane	5.471	56	26854	17.25	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	27938	17.67	ug/l	95
36) 1,1-Dichloropropene	5.696	75	25167	18.40	ug/l	96
37) Ethyl Acetate	4.727	43	31097	17.89	ug/l	95
38) Carbon Tetrachloride	5.678	117	22251	19.02	ug/l	94
39) Methylcyclohexane	7.385	83	26709	18.72	ug/l	94
40) Benzene	6.044	78	76683	18.85	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022702.D
 Acq On : 15 Jun 2021 16:30
 Operator : JC/MD
 Sample : VX0615WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0615WBS01

Quant Time: Jun 16 05:26:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	17931	19.25	ug/l	91
42) 1,2-Dichloroethane	6.092	62	27998	18.71	ug/l	99
43) Isopropyl Acetate	6.348	43	52206	19.02	ug/l #	93
44) Trichloroethene	7.129	130	18217	19.07	ug/l	91
45) 1,2-Dichloropropane	7.434	63	20644	18.80	ug/l	98
46) Dibromomethane	7.586	93	13140	18.64	ug/l #	84
47) Bromodichloromethane	7.824	83	24450	18.78	ug/l	97
48) Methyl methacrylate	7.696	41	24104	18.29	ug/l	87
49) 1,4-Dioxane	7.665	88	9707	387.00	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	166990	95.83	ug/l	90
52) Toluene	8.720	92	47749	18.85	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	26994	18.04	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	30553	18.37	ug/l	90
55) 1,1,2-Trichloroethane	9.153	97	19638	18.95	ug/l	96
56) Ethyl methacrylate	9.122	69	30991	18.07	ug/l #	83
57) 1,3-Dichloropropane	9.311	76	34501	19.14	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.244	63	84936	90.92	ug/l	95
59) 2-Hexanone	9.433	43	126713	95.17	ug/l	88
60) Dibromochloromethane	9.525	129	16167	18.05	ug/l	98
61) 1,2-Dibromoethane	9.610	107	19726	19.07	ug/l	98
64) Tetrachloroethene	9.275	164	16096	20.15	ug/l	91
65) Chlorobenzene	10.079	112	49283	18.72	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	16137	18.14	ug/l	98
67) Ethyl Benzene	10.195	91	92733	19.14	ug/l	97
68) m/p-Xylenes	10.305	106	67809	38.63	ug/l	96
69) o-Xylene	10.646	106	32908	19.35	ug/l	97
70) Styrene	10.659	104	55479	18.60	ug/l	97
71) Bromoform	10.805	173	9173	18.03	ug/l #	95
73) Isopropylbenzene	10.963	105	88404	19.81	ug/l	98
74) N-amyl acetate	10.848	43	44391	18.50	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	11.213	83	31694	19.97	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	34157	23.70	ug/l	89
77) Bromobenzene	11.201	156	18211	19.10	ug/l	83
78) n-propylbenzene	11.305	91	104627	19.41	ug/l	97
79) 2-Chlorotoluene	11.366	91	63877	19.58	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	74195	19.96	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	7903	18.49	ug/l #	80
82) 4-Chlorotoluene	11.457	91	71701	19.21	ug/l	96
83) tert-Butylbenzene	11.719	119	69171	19.81	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	74791	19.84	ug/l	97
85) sec-Butylbenzene	11.896	105	90903	19.93	ug/l	97
86) p-Isopropyltoluene	12.012	119	74725	19.85	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	35633	19.69	ug/l	95
88) 1,4-Dichlorobenzene	12.043	146	35373	18.99	ug/l	95
89) n-Butylbenzene	12.335	91	71909	19.95	ug/l	96
90) Hexachloroethane	12.542	117	9858	18.27	ug/l	74
91) 1,2-Dichlorobenzene	12.335	146	34861	19.77	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5633	17.80	ug/l	79
93) 1,2,4-Trichlorobenzene	13.591	180	20216	19.11	ug/l	94
94) Hexachlorobutadiene	13.725	225	7790	20.54	ug/l	97
95) Naphthalene	13.780	128	81869	18.68	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	20507	19.89	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
Data File : VX022702.D
Acq On : 15 Jun 2021 16:30
Operator : JC/MD
Sample : VX0615WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01

Quant Time: Jun 16 05:26:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022702.D
 Acq On : 15 Jun 2021 16:30
 Operator : JC/MD
 Sample : VX0615WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0615WBS01

Quant Time: Jun 16 05:26:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:168 Resp: 71762
Ion Ratio Lower Upper
168 100
99 65.6 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 18.31 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.007 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 85 Resp: 15419
Ion Ratio Lower Upper
85 100
87 31.1 16.4 49.2









#3
Chloromethane
Concen: 16.42 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 50 Resp: 16706
Ion Ratio Lower Upper
50 100
52 32.7 26.4 39.6



#4
Vinyl Chloride
Concen: 16.69 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 62 Resp: 16690
Ion Ratio Lower Upper
62 100
64 31.5 25.7 38.5









#5
Bromomethane
Concen: 16.72 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 94 Resp: 9093
Ion Ratio Lower Upper
94 100
96 97.1 76.4 114.6



#6
Chloroethane
Concen: 16.86 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 64 Resp: 10205
Ion Ratio Lower Upper
64 100
66 37.8 25.4 38.0









#7
Trichlorofluoromethane
Concen: 17.64 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:101 Resp: 25405
Ion Ratio Lower Upper
101 100
103 63.4 53.2 79.8



#8
Diethyl Ether
Concen: 17.75 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 74 Resp: 10323
Ion Ratio Lower Upper
74 100
45 105.0 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.06 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:101 Resp: 15866
Ion Ratio Lower Upper
101 100
85 48.8 34.9 52.3
151 64.6 61.4 92.2



#10
Methyl Iodide
Concen: 13.21 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:142 Resp: 13012
Ion Ratio Lower Upper
142 100
127 50.0 33.3 49.9#
141 14.6 11.4 17.0









#11

Tert butyl alcohol
Concen: 82.35 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01




Tgt Ion: 59 Resp: 24608
Ion Ratio Lower Upper
59 100
57 10.2 7.8 11.8




#12

1,1-Dichloroethene
Concen: 17.99 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 96 Resp: 15673
Ion Ratio Lower Upper
96 100
61 162.5 124.6 187.0
98 62.9 50.6 76.0









#13
Acrolein
Concen: 101.41 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 56 Resp: 12700
Ion Ratio Lower Upper
56 100
55 70.3 56.4 84.6



#14
Allyl chloride
Concen: 17.26 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 41 Resp: 29075
Ion Ratio Lower Upper
41 100
39 68.9 51.6 77.4
76 31.6 30.5 45.7









#15
Acrylonitrile
Concen: 89.95 ug/l
RT: 3.068 min Scan# 325
Delta R.T. -0.007 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:	53	Resp:	52789
Ion	Ratio	Lower	Upper
53	100		
52	81.5	65.3	97.9
51	35.5	28.2	42.2



#16
Acetone
Concen: 83.35 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:	43	Resp:	48166
Ion	Ratio	Lower	Upper
43	100		
58	33.3	27.8	41.8









#17
Carbon Disulfide
Concen: 15.94 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 76 Resp: 31252
Ion Ratio Lower Upper
76 100
78 10.0 7.2 10.8



#18
Methyl Acetate
Concen: 17.10 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 43 Resp: 29749
Ion Ratio Lower Upper
43 100
74 24.0 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 17.87 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 73 Resp: 54028
Ion Ratio Lower Upper
73 100
57 21.0 17.1 25.7



#20
Methylene Chloride
Concen: 18.09 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 84 Resp: 20379
Ion Ratio Lower Upper
84 100
49 128.1 88.6 132.8
51 39.5 27.0 40.6
86 56.7 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 16.88 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 96 Resp: 16299
Ion Ratio Lower Upper
96 100
61 151.6 107.7 161.5
98 66.6 50.4 75.6



#22
Diisopropyl ether
Concen: 17.78 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 45 Resp: 59252
Ion Ratio Lower Upper
45 100
43 71.2 58.0 87.0
87 25.5 25.4 38.0
59 11.1 10.2 15.2









#23
 Vinyl Acetate
 Concen: 89.21 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.000 min
 Lab File: VX022702.D
 Acq: 15 Jun 2021 16:30




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0615WBS01


Tgt Ion: 43 Resp: 274062
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 17.79 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX022702.D
 Acq: 15 Jun 2021 16:30

Tgt Ion: 63 Resp: 34415
 Ion Ratio Lower Upper
 63 100
 98 6.2 3.8 11.2
 100 4.3 2.5 7.5





#25

2-Butanone

Concen: 86.53 ug/l

RT: 4.568 min Scan# 571

Delta R.T. 0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion: 43 Resp: 81097

Ion Ratio Lower Upper

43 100





72 23.5 23.7 35.5#

Instrument :

MSVOA_X

ClientSampleId :

VX0615WBS01



#26

2,2-Dichloropropane

Concen: 18.93 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX022702.D




Acq: 15 Jun 2021 16:30


Tgt Ion: 77 Resp: 27930

Ion Ratio Lower Upper

77 100

97 22.2 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 17.62 ug/l

RT: 4.495 min Scan# 559





Delta R.T. 0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion: 96 Resp: 20731

Ion	Ratio	Lower	Upper
96	100		
61	148.3	0.0	285.6
98	64.4	0.0	129.6



#28

Bromochloromethane

Concen: 18.58 ug/l

RT: 4.910 min Scan# 627




Delta R.T. 0.006 min


Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion: 49 Resp: 17375

Ion	Ratio	Lower	Upper
49	100		
129	1.1	0.0	4.0
130	61.1	68.2	102.2#





#29

Tetrahydrofuran

Concen: 87.77 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion: 42 Resp: 47899

Ion Ratio Lower Upper

42 100

72 43.0 41.7 62.5





71 40.6 38.4 57.6

Instrument :

MSVOA_X

ClientSampleId :

VX0615WBS01



#30

Chloroform

Concen: 17.92 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX022702.D




Acq: 15 Jun 2021 16:30


Tgt Ion: 83 Resp: 34559

Ion Ratio Lower Upper

83 100

85 69.1 52.0 78.0









#31
Cyclohexane
Concen: 17.25 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 56 Resp: 26854
Ion Ratio Lower Upper
56 100
69 28.2 25.3 37.9
84 89.5 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 17.67 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 97 Resp: 27938
Ion Ratio Lower Upper
97 100
99 62.5 51.5 77.3
61 47.8 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 45.55 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX022702.D





Acq: 15 Jun 2021 16:30

Tgt Ion: 65 Resp: 58782

Ion Ratio Lower Upper

65 100

67 54.2 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30




Tgt Ion: 114 Resp: 137264


Ion Ratio Lower Upper

114 100

63 21.7 0.0 40.8

88 16.3 0.0 32.8





#35

Dibromofluoromethane

Concen: 48.93 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion:113 Resp: 42848

Ion Ratio Lower Upper

113 100

111 103.9 81.9 122.9





192 15.2 16.7 25.1#

Instrument :

MSVOA_X

ClientSampleId :

VX0615WBS01



#36

1,1-Dichloropropene

Concen: 18.40 ug/l

RT: 5.696 min Scan# 756

Delta R.T. -0.006 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30




Tgt Ion: 75 Resp: 25167


Ion Ratio Lower Upper

75 100

110 32.7 18.1 54.4

77 31.3 25.3 37.9





#37

Ethyl Acetate

Concen: 17.89 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30





Instrument :

MSVOA_X

ClientSampleId :

VX0615WBS01

Tgt Ion:	43	Resp:	31097
Ion	Ratio	Lower	Upper
43	100		
61	14.2	12.8	19.2
70	11.0	10.5	15.7



#38

Carbon Tetrachloride

Concen: 19.02 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion:	117	Resp:	22251
Ion	Ratio	Lower	Upper
117	100		
119	90.9	76.2	114.2
121	25.9	24.8	37.2





#39

Methylcyclohexane

Concen: 18.72 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion: 83 Resp: 26709





Ion	Ratio	Lower	Upper
83	100		
55	80.6	58.9	88.3
98	48.8	38.1	57.1

Instrument :

MSVOA_X

ClientSampleId :

VX0615WBS01



#40

Benzene

Concen: 18.85 ug/l

RT: 6.044 min Scan# 813




Delta R.T. -0.006 min


Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion: 78 Resp: 76683

Ion	Ratio	Lower	Upper
78	100		
77	22.6	18.9	28.3









#41
Methacrylonitrile
Concen: 19.25 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:	41	Resp:	17931
Ion	Ratio	Lower	Upper
41	100		
39	52.5	42.5	63.7
67	68.1	65.0	97.6
52	27.7	25.3	37.9



#42
1,2-Dichloroethane
Concen: 18.71 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:	62	Resp:	27998
Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	20.2









#43
Isopropyl Acetate
Concen: 19.02 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 43	Resp: 52206
Ion Ratio	Lower Upper
43 100	
61 19.6	18.2 27.4
87 11.9	12.5 18.7#



#44
Trichloroethene
Concen: 19.07 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 130	Resp: 18217
Ion Ratio	Lower Upper
130 100	
95 107.6	0.0 198.0









#45
1,2-Dichloropropane
Concen: 18.80 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampled :
VX0615WBS01


Tgt Ion: 63 Resp: 20644
Ion Ratio Lower Upper
63 100
65 30.8 25.7 38.5



#46
Dibromomethane
Concen: 18.64 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 93 Resp: 13140
Ion Ratio Lower Upper
93 100
95 85.2 66.8 100.2
174 76.7 83.1 124.7#









#47
Bromodichloromethane
Concen: 18.78 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 83 Resp: 24450
Ion Ratio Lower Upper
83 100
85 62.3 51.4 77.0
127 7.2 6.8 10.2



#48
Methyl methacrylate
Concen: 18.29 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 41 Resp: 24104
Ion Ratio Lower Upper
41 100
69 81.0 76.0 114.0
39 63.4 44.6 67.0









#49
1,4-Dioxane
Concen: 387.00 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampled :
VX0615WBS01


Tgt Ion: 88 Resp: 9707
Ion Ratio Lower Upper
88 100
43 34.7 23.2 34.8
58 69.0 55.0 82.4



#50
Toluene-d8
Concen: 49.75 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 98 Resp: 171772
Ion Ratio Lower Upper
98 100
100 66.0 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 95.83 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 43 Resp: 166990
Ion Ratio Lower Upper
43 100
58 37.8 35.4 53.0



#52
Toluene
Concen: 18.85 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 92 Resp: 47749
Ion Ratio Lower Upper
92 100
91 170.7 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 18.04 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampled :
VX0615WBS01


Tgt Ion: 75 Resp: 26994
Ion Ratio Lower Upper
75 100
77 32.4 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 18.37 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 75 Resp: 30553
Ion Ratio Lower Upper
75 100
77 29.2 25.7 38.5
39 52.3 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 18.95 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:	97	Resp:	19638
Ion	Ratio	Lower	Upper
97	100		
83	90.7	69.3	103.9
85	59.5	44.3	66.5
99	61.7	50.3	75.5



#56
Ethyl methacrylate
Concen: 18.07 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:	69	Resp:	30991
Ion	Ratio	Lower	Upper
69	100		
41	69.0	46.8	70.2
39	42.7	24.7	37.1









#57

1,3-Dichloropropane
Concen: 19.14 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Instrument :
MSVOA_X
ClientSampled :
VX0615WBS01




Tgt Ion: 76 Resp: 34501
Ion Ratio Lower Upper
76 100
78 33.8 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 90.92 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 63 Resp: 84936
Ion Ratio Lower Upper
63 100
106 26.9 23.5 35.3





#59

2-Hexanone

Concen: 95.17 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX022702.D





Acq: 15 Jun 2021 16:30

Tgt Ion: 43 Resp: 126713

Ion Ratio Lower Upper

43 100

58 52.7 30.9 92.8



#60

Dibromochloromethane

Concen: 18.05 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX022702.D




Acq: 15 Jun 2021 16:30


Tgt Ion:129 Resp: 16167

Ion Ratio Lower Upper

129 100

127 78.2 38.4 115.2









#61
1,2-Dibromoethane
Concen: 19.07 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:107 Resp: 19726
Ion Ratio Lower Upper
107 100
109 92.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.61 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 95 Resp: 63463
Ion Ratio Lower Upper
95 100
174 61.6 0.0 154.6
176 60.4 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:117 Resp: 126306
Ion Ratio Lower Upper
117 100
82 62.2 47.4 71.2
119 32.4 26.6 39.8



#64
Tetrachloroethene
Concen: 20.15 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:164 Resp: 16096
Ion Ratio Lower Upper
164 100
166 124.5 103.4 155.2
129 102.3 70.8 106.2
131 100.6 72.5 108.7









#65
Chlorobenzene
Concen: 18.72 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:112 Resp: 49283
Ion Ratio Lower Upper
112 100
114 30.6 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 18.14 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:131 Resp: 16137
Ion Ratio Lower Upper
131 100
133 95.3 47.7 143.1
119 68.4 32.1 96.5









#67
Ethyl Benzene
Concen: 19.14 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 91 Resp: 92733
Ion Ratio Lower Upper
91 100
106 29.1 24.6 37.0



#68
m/p-Xylenes
Concen: 38.63 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 106 Resp: 67809
Ion Ratio Lower Upper
106 100
91 210.0 162.8 244.2









#69
o-Xylene
Concen: 19.35 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:106 Resp: 32908
Ion Ratio Lower Upper
106 100
91 221.3 108.1 324.2



#70
Styrene
Concen: 18.60 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:104 Resp: 55479
Ion Ratio Lower Upper
104 100
78 52.7 40.0 60.0
103 53.2 43.4 65.0





#71

Bromoform

Concen: 18.03 ug/l

RT: 10.805 min Scan# 1594





Delta R.T. -0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion:173	Resp:	9173
Ion Ratio	Lower	Upper
173	100	
175	44.8	23.9 71.8
254	0.0	0.1 0.1#

Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. -0.000 min

Lab File: VX022702.D

Acq: 15 Jun 2021 16:30

Tgt Ion:152	Resp:	54349
Ion Ratio	Lower	Upper
152	100	
115	78.6	41.1 123.3
150	169.7	0.0 349.0









#73
Isopropylbenzene
Concen: 19.81 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:105 Resp: 88404
Ion Ratio Lower Upper
105 100
120 25.4 13.2 39.6



#74
N-ethyl acetate
Concen: 18.50 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 43 Resp: 44391
Ion Ratio Lower Upper
43 100
70 38.6 39.0 58.6#
55 22.4 21.5 32.3
61 23.6 22.1 33.1









#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.97 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX022702.D
 Acq: 15 Jun 2021 16:30




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0615WBS01


Tgt Ion: 83 Resp: 31694
 Ion Ratio Lower Upper
 83 100
 131 9.0 5.1 15.2
 85 62.8 31.8 95.4



#76
 1,2,3-Trichloropropane
 Concen: 23.70 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX022702.D
 Acq: 15 Jun 2021 16:30

Tgt Ion: 75 Resp: 34157
 Ion Ratio Lower Upper
 75 100
 77 31.9 19.1 57.5









#77
Bromobenzene
Concen: 19.10 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:156 Resp: 18211
Ion Ratio Lower Upper
156 100
77 200.1 82.8 248.6
158 96.1 49.1 147.4



#78
n-propylbenzene
Concen: 19.41 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 91 Resp: 104627
Ion Ratio Lower Upper
91 100
120 21.8 11.5 34.5









#79
2-Chlorotoluene
Concen: 19.58 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 91 Resp: 63877
Ion Ratio Lower Upper
91 100
126 31.0 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 19.96 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 105 Resp: 74195
Ion Ratio Lower Upper
105 100
120 46.3 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 18.49 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 75 Resp: 7903
Ion Ratio Lower Upper
75 100
53 119.8 74.5 111.7#
89 44.0 37.9 56.9



#82
4-Chlorotoluene
Concen: 19.21 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 91 Resp: 71701
Ion Ratio Lower Upper
91 100
126 27.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 19.81 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:119 Resp: 69171
Ion Ratio Lower Upper
119 100
91 65.4 27.9 83.7
134 23.8 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 19.84 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:105 Resp: 74791
Ion Ratio Lower Upper
105 100
120 42.9 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 19.93 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022702.D
 Acq: 15 Jun 2021 16:30




Instrument :
 MSVOA_X
 ClientSampled :
 VX0615WBS01


Tgt Ion:105 Resp: 90903
 Ion Ratio Lower Upper
 105 100
 134 18.7 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 19.85 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022702.D
 Acq: 15 Jun 2021 16:30

Tgt Ion:119 Resp: 74725
 Ion Ratio Lower Upper
 119 100
 134 26.4 12.9 38.7
 91 26.0 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 19.69 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampled :
VX0615WBS01


Tgt Ion:146 Resp: 35633
Ion Ratio Lower Upper
146 100
111 45.5 20.2 60.5
148 61.4 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 18.99 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:146 Resp: 35373
Ion Ratio Lower Upper
146 100
111 46.1 19.6 58.8
148 63.9 31.6 95.0









#89
n-Butylbenzene
Concen: 19.95 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion: 91	Resp:	71909
Ion	Ratio	Lower Upper
91	100	
92	52.6	27.4 82.0
134	22.9	13.2 39.5



#90
Hexachloroethane
Concen: 18.27 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 117	Resp:	9858
Ion	Ratio	Lower Upper
117	100	
201	57.8	40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 19.77 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:146 Resp: 34861
Ion Ratio Lower Upper
146 100
111 47.6 20.8 62.2
148 63.7 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.80 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion: 75 Resp: 5633
Ion Ratio Lower Upper
75 100
155 70.9 45.3 135.9
157 92.1 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 19.11 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30




Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01


Tgt Ion:180	Resp:	20216
Ion Ratio	Lower	Upper
180	100	
182	91.4	47.7 143.1
145	35.5	15.0 45.0



#94
Hexachlorobutadiene
Concen: 20.54 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.006 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:225	Resp:	7790
Ion Ratio	Lower	Upper
225	100	
223	61.1	31.1 93.2
227	59.5	31.6 95.0









#95
Naphthalene
Concen: 18.68 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Instrument :
MSVOA_X
ClientSampleId :
VX0615WBS01

Tgt Ion:128 Resp: 81869
Ion Ratio Lower Upper
128 100
127 12.1 10.3 15.5
129 11.2 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.89 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022702.D
Acq: 15 Jun 2021 16:30

Tgt Ion:180 Resp: 20507
Ion Ratio Lower Upper
180 100
182 93.0 47.6 142.9
145 36.1 16.1 48.3

