

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022705.D
 Acq On : 15 Jun 2021 17:44
 Operator : JC/MD
 Sample : M2712-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Jun 16 05:27:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56358	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.775	114	112935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47503	46.866	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.740%
35) Dibromofluoromethane	5.397	113	34520	47.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.820%
50) Toluene-d8	8.653	98	136644	48.106	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.220%
62) 4-Bromofluorobenzene	11.085	95	46338	43.135	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.260%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	482	0.704	ug/l	# 88
16) Acetone	2.392	43	714	1.573	ug/l	93
27) cis-1,2-Dichloroethene	4.501	96	937	1.014	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
Data File : VX022705.D
Acq On : 15 Jun 2021 17:44
Operator : JC/MD
Sample : M2712-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Jun 16 05:27:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

