



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061521\
 Data File : VX022698.D
 Acq On : 15 Jun 2021 14:25
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Title : SW846 8260
 Last Update : Thu Jun 10 12:20:00 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	4898	PASS
75	95	30	60	57.0	12817	PASS
95	95	100	100	100.0	22503	PASS
96	95	5	9	5.0	1128	PASS
173	174	0.00	2	0.5	68	PASS
174	95	50	100	57.5	12930	PASS
175	174	5	9	7.5	976	PASS
176	174	95	101	97.2	12571	PASS
177	176	5	9	6.5	822	PASS