

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
 Data File : VX029504.D
 Acq On : 15 Jun 2022 13:20
 Operator : JC/MD
 Sample : VX0615MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0615MBL01

Quant Time: Jun 15 15:37:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	179252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	300795	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170822	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149527	54.974	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.940%
35) Dibromofluoromethane	5.391	113	121893	47.667	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.340%
50) Toluene-d8	8.653	98	453283	50.120	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.240%
62) 4-Bromofluorobenzene	11.085	95	170803	48.260	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
Data File : VX029504.D
Acq On : 15 Jun 2022 13:20
Operator : JC/MD
Sample : VX0615MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0615MBL01

Quant Time: Jun 15 15:37:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

