

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
 Data File : VX029512.D
 Acq On : 15 Jun 2022 17:54
 Operator : JC/MD
 Sample : N3304-05
 Misc : 4.72g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-1A-6

Quant Time: Jun 15 18:49:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	278947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	367592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	167717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	232077	54.825	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 109.640%			
35) Dibromofluoromethane	5.391	113	192678	49.814	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.620%			
50) Toluene-d8	8.653	98	775735	56.707	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 113.420%			
62) 4-Bromofluorobenzene	11.086	95	292341	54.609	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 109.220%			
Target Compounds					Qvalue	
31) Cyclohexane	5.458	56	203826	26.954	ug/l	94
39) Methylcyclohexane	7.379	83	1099008	139.645	ug/l	95
43) Isopropyl Acetate	6.349	43	16408	1.656	ug/l #	58
44) Trichloroethene	7.129	130	7259	1.402	ug/l	90
52) Toluene	8.720	92	9046	0.741	ug/l	96
67) Ethyl Benzene	10.195	91	71472	4.013	ug/l	96
68) m/p-Xylenes	10.305	106	74054	10.748	ug/l	94
69) o-Xylene	10.647	106	49358	7.372	ug/l	96
73) Isopropylbenzene	10.964	105	89121	6.750	ug/l	100
78) n-propylbenzene	11.305	91	139117	9.285	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	227988	20.335	ug/l	100
83) tert-Butylbenzene	11.720	119	13858	1.274	ug/l	83
84) 1,2,4-Trimethylbenzene	11.756	105	533430	48.204	ug/l	100
85) sec-Butylbenzene	11.896	105	111455	8.343	ug/l #	76
86) p-Isopropyltoluene	12.012	119	138080	12.330	ug/l	96
89) n-Butylbenzene	12.335	91	140511	14.882	ug/l #	34

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
Data File : VX029512.D
Acq On : 15 Jun 2022 17:54
Operator : JC/MD
Sample : N3304-05
Misc : 4.72g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-1A-6

Quant Time: Jun 15 18:49:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

