

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
 Data File : VX029513.D
 Acq On : 15 Jun 2022 18:18
 Operator : JC/MD
 Sample : N3304-06
 Misc : 5.15g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-1A-6-G

Quant Time: Jun 15 19:22:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	291714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	467206	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	213867	48.096	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.200%
35) Dibromofluoromethane	5.391	113	177939	44.799	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.600%
50) Toluene-d8	8.653	98	710159	50.554	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.100%
62) 4-Bromofluorobenzene	11.085	95	262124	47.682	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.360%
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.739	43	12341	1.974	ug/l	96
31) Cyclohexane	5.465	56	11771	1.488	ug/l	92
39) Methylcyclohexane	7.373	83	151485	18.745	ug/l	90
44) Trichloroethene	7.129	130	6951	1.307	ug/l	96
65) Chlorobenzene	10.080	112	8063	0.765	ug/l	89
84) 1,2,4-Trimethylbenzene	11.756	105	15565	1.389	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	100736	16.526	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
Data File : VX029513.D
Acq On : 15 Jun 2022 18:18
Operator : JC/MD
Sample : N3304-06
Misc : 5.15g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-1A-6-G

Quant Time: Jun 15 19:22:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

