

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
 Data File : VX029516.D
 Acq On : 15 Jun 2022 19:28
 Operator : JC/MD
 Sample : N3293-07ME
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40808ME

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/16/2022
 Supervised By :Semsettin Yesilyurt 06/16/2022

Quant Time: Jun 16 04:16:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	282530	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	461361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	392967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155447	35.641	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	71.280%		
35) Dibromofluoromethane	5.397	113	123740	31.548	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	63.100%#		
50) Toluene-d8	8.653	98	524673	37.823	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	75.640%		
62) 4-Bromofluorobenzene	11.085	95	187662	34.570	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	69.140%		
Target Compounds						
6) Chloroethane	1.709	64	153	Below Cal	#	69
11) Tert butyl alcohol	2.953	59	12312	Below Cal	#	74
16) Acetone	2.435	43	9375	4.102	ug/l	95
25) 2-Butanone	4.617	43	9722m	2.689	ug/l	
41) Methacrylonitrile	4.897	41	22	Below Cal	#	18
44) Trichloroethene	7.129	130	6355	1.210	ug/l	94
48) Methyl methacrylate	7.720	41	1899	0.376	ug/l	# 37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
Data File : VX029516.D
Acq On : 15 Jun 2022 19:28
Operator : JC/MD
Sample : N3293-07ME
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40808ME

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/16/2022
Supervised By :Semsettin Yesilyurt 06/16/2022

Quant Time: Jun 16 04:16:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

