

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
 Data File : VX029502.D
 Acq On : 15 Jun 2022 10:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 15 15:37:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.794	0.760	4.3	85	0.00
3 P	Chloromethane	0.847	0.928	-9.6	103	0.00
4 C	Vinyl Chloride	0.879	0.848	3.5#	86	0.00
5 T	Bromomethane	0.471	0.380	19.3	76	0.00
6 T	Chloroethane	0.446	0.512	-14.8	115	0.00
7 T	Trichlorofluoromethane	1.281	1.222	4.6	85	0.00
8 T	Diethyl Ether	0.499	0.487	2.4	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.772	0.767	0.6	91	0.00
10 T	Methyl Iodide	0.879	0.707	19.6	66	0.00
11 T	Tert butyl alcohol	0.244	0.165	32.4#	69	0.00
12 CM	1,1-Dichloroethene	0.752	0.729	3.1#	87	0.00
13 T	Acrolein	0.092	0.080	13.0	82	0.00
14 T	Allyl chloride	1.207	1.177	2.5	87	0.00
15 T	Acrylonitrile	0.476	0.443	6.9	83	0.00
16 T	Acetone	0.404	0.475	-17.6	109	0.00
17 T	Carbon Disulfide	1.850	1.715	7.3	83	0.00
18 T	Methyl Acetate	1.072	1.022	4.7	87	0.00
19 T	Methyl tert-butyl Ether	2.595	2.647	-2.0	90	0.00
20 T	Methylene Chloride	0.903	0.842	6.8	87	0.00
21 T	trans-1,2-Dichloroethene	0.844	0.814	3.6	88	0.00
22 T	Diisopropyl ether	2.435	2.468	-1.4	88	0.00
23 T	Vinyl Acetate	2.019	2.068	-2.4	85	0.00
24 P	1,1-Dichloroethane	1.469	1.448	1.4	87	0.00
25 T	2-Butanone	0.640	0.642	-0.3	88	0.00
26 T	2,2-Dichloropropane	1.047	1.169	-11.7	99	0.00
27 T	cis-1,2-Dichloroethene	0.998	0.957	4.1	88	0.00
28 T	Bromochloromethane	0.577	0.604	-4.7	94	0.00
29 T	Tetrahydrofuran	0.417	0.396	5.0	82	0.00
30 C	Chloroform	1.585	1.545	2.5#	88	0.00
31 T	Cyclohexane	1.355	1.329	1.9	88	0.00
32 T	1,1,1-Trichloroethane	1.366	1.353	1.0	87	0.00
33 S	1,2-Dichloroethane-d4	0.818	0.719	12.1	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.425	0.397	6.6	92	0.00
36 T	1,1-Dichloropropene	0.702	0.705	-0.4	87	0.00
37 T	Ethyl Acetate	0.725	0.717	1.1	83	0.00
38 T	Carbon Tetrachloride	0.718	0.685	4.6	83	0.00
39 T	Methylcyclohexane	0.865	0.929	-7.4	93	0.00
40 TM	Benzene	2.090	2.063	1.3	86	0.00
41 T	Methacrylonitrile	0.360	0.403	-11.9	87	0.00
42 TM	1,2-Dichloroethane	0.726	0.710	2.2	85	0.00
43 T	Isopropyl Acetate	1.089	1.112	-2.1	87	0.00
44 TM	Trichloroethene	0.569	0.587	-3.2	93	0.00
45 C	1,2-Dichloropropane	0.514	0.521	-1.4#	89	0.00
46 T	Dibromomethane	0.372	0.372	0.0	87	0.00
47 T	Bromodichloromethane	0.703	0.720	-2.4	87	0.00
48 T	Methyl methacrylate	0.548	0.553	-0.9	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
 Data File : VX029502.D
 Acq On : 15 Jun 2022 10:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 15 15:37:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.014	0.012	14.3	72	-0.01
50 S	Toluene-d8	1.503	1.340	10.8	89	0.00
51 T	4-Methyl-2-Pentanone	0.713	0.721	-1.1	84	0.00
52 CM	Toluene	1.342	1.369	-2.0#	87	0.00
53 T	t-1,3-Dichloropropene	0.683	0.767	-12.3	90	0.00
54 T	cis-1,3-Dichloropropene	0.789	0.846	-7.2	88	0.00
55 T	1,1,2-Trichloroethane	0.543	0.543	0.0	86	0.00
56 T	Ethyl methacrylate	0.790	0.856	-8.4	87	0.00
57 T	1,3-Dichloropropane	0.915	0.913	0.2	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.384	0.377	1.8	82	0.00
59 T	2-Hexanone	0.542	0.560	-3.3	84	0.00
60 T	Dibromochloromethane	0.525	0.538	-2.5	85	0.00
61 T	1,2-Dibromoethane	0.579	0.578	0.2	85	0.00
62 S	4-Bromofluorobenzene	0.588	0.559	4.9	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.474	0.505	-6.5	91	0.00
65 PM	Chlorobenzene	1.378	1.402	-1.7	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.480	0.498	-3.8	85	0.00
67 C	Ethyl Benzene	2.422	2.554	-5.5#	88	0.00
68 T	m/p-Xylenes	0.937	1.000	-6.7	88	0.00
69 T	o-Xylene	0.911	0.986	-8.2	89	0.00
70 T	Styrene	1.520	1.652	-8.7	89	0.00
71 P	Bromoform	0.341	0.354	-3.8	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.936	4.064	-3.3	89	0.00
74 T	N-amyl acetate	1.333	1.391	-4.4	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.328	1.270	4.4	85	0.00
76 T	1,2,3-Trichloropropane	1.168	1.166	0.2	86	0.00
77 T	Bromobenzene	0.924	0.919	0.5	88	0.00
78 T	n-propylbenzene	4.467	4.709	-5.4	89	0.00
79 T	2-Chlorotoluene	2.751	2.821	-2.5	89	0.00
80 T	1,3,5-Trimethylbenzene	3.342	3.467	-3.7	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.349	0.345	1.1	81	0.00
82 T	4-Chlorotoluene	3.175	3.229	-1.7	88	0.00
83 T	tert-Butylbenzene	3.243	3.375	-4.1	89	0.00
84 T	1,2,4-Trimethylbenzene	3.299	3.451	-4.6	89	0.00
85 T	sec-Butylbenzene	3.982	4.336	-8.9	92	0.00
86 T	p-Isopropyltoluene	3.339	3.664	-9.7	92	0.00
87 T	1,3-Dichlorobenzene	1.764	1.801	-2.1	89	0.00
88 T	1,4-Dichlorobenzene	1.794	1.790	0.2	89	0.00
89 T	n-Butylbenzene	2.815	3.226	-14.6	96	0.00
90 T	Hexachloroethane	0.531	0.500	5.8	81	0.00
91 T	1,2-Dichlorobenzene	1.749	1.766	-1.0	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.276	0.275	0.4	84	0.00
93 T	1,2,4-Trichlorobenzene	1.032	1.153	-11.7	96	0.00
94 T	Hexachlorobutadiene	0.429	0.488	-13.8	101	0.00
95 T	Naphthalene	3.643	4.030	-10.6	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
Data File : VX029502.D
Acq On : 15 Jun 2022 10:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 15 15:37:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.029	1.131	-9.9	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6