



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061522\
Data File : VX029501.D
Acq On : 15 Jun 2022 08:58
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Title : SW846 8260
Last Update : Wed Jun 08 03:00:38 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	11280	PASS
75	95	30	60	49.1	33557	PASS
95	95	100	100	100.0	68352	PASS
96	95	5	9	6.9	4691	PASS
173	174	0.00	2	0.9	430	PASS
174	95	50	100	67.4	46045	PASS
175	174	5	9	7.1	3285	PASS
176	174	95	101	95.6	44027	PASS
177	176	5	9	6.5	2877	PASS