

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061523\
 Data File : VX036171.D
 Acq On : 15 Jun 2023 15:52
 Operator : JC/MD
 Sample : 03278-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20230614

Quant Time: Jun 16 07:56:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	38790	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	205919	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	180959	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	106430	33.022	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.067%#
60) 4-Bromofluorobenzene	11.079	95	88155	29.032	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.767%
63) Toluene-d8	8.647	98	232322	29.041	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061523\
Data File : VX036171.D
Acq On : 15 Jun 2023 15:52
Operator : JC/MD
Sample : 03278-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20230614

Quant Time: Jun 16 07:56:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

