

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061523\
 Data File : VX036173.D
 Acq On : 15 Jun 2023 16:39
 Operator : JC/MD
 Sample : 03251-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01

Quant Time: Jun 16 07:57:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	36570	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.763	114	221704	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	205679	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	118563	39.019	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	130.067%#
60) 4-Bromofluorobenzene	11.079	95	93721	27.156	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	90.533%
63) Toluene-d8	8.647	98	249500	27.440	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	91.467%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	1098	2.827	ug/l #	49
25) Chloroform	5.099	83	3267	0.761	ug/l #	83
37) Trichloroethene	7.129	130	1267	0.518	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061523\
Data File : VX036173.D
Acq On : 15 Jun 2023 16:39
Operator : JC/MD
Sample : 03251-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01

Quant Time: Jun 16 07:57:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

