

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061523\
 Data File : VX036175.D
 Acq On : 15 Jun 2023 17:25
 Operator : JC/MD
 Sample : 03251-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01RE

Quant Time: Jun 16 07:58:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	37480	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	222850	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	201348	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	118293	37.985	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	126.633%#
60) 4-Bromofluorobenzene	11.079	95	93191	27.583	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.933%
63) Toluene-d8	8.647	98	248214	27.885	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	92.967%
Target Compounds						
					Qvalue	
25) Chloroform	5.092	83	3266	0.742	ug/l	95
37) Trichloroethene	7.123	130	1459	0.593	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061523\
Data File : VX036175.D
Acq On : 15 Jun 2023 17:25
Operator : JC/MD
Sample : 03251-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01RE

Quant Time: Jun 16 07:58:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

