

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061620\
 Data File : VX016778.D
 Acq On : 16 Jun 2020 14:44
 Operator : JC/SP
 Sample : L3003-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-TB-2020-2

Quant Time: Jun 17 04:29:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	252657	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	395369	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	401373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	143648	50.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.10%	
35) Dibromofluoromethane	5.46	113	118968	48.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.54%	
50) Toluene-d8	8.70	98	482925	53.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.38%	
62) 4-Bromofluorobenzene	11.12	95	197885	57.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061620\
Data File : VX016778.D
Acq On : 16 Jun 2020 14:44
Operator : JC/SP
Sample : L3003-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-TB-2020-2

Quant Time: Jun 17 04:29:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

