

Data File : VX016779.D
Acq On : 16 Jun 2020 15:07
Operator : JC/SP
Sample : L2993-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW039-200610

Quant Time: Jun 17 04:31:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	241098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	378799	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	392450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195168	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	138662	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
35) Dibromofluoromethane	5.47	113	115251	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
50) Toluene-d8	8.70	98	470766	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
62) 4-Bromofluorobenzene	11.12	95	189243	57.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.82%	
Target Compounds						
16) Acetone	2.42	43	2993	2.433	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX016779.D
Acq On : 16 Jun 2020 15:07
Operator : JC/SP
Sample : L2993-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW039-200610

Quant Time: Jun 17 04:31:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

