

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061620\
 Data File : VX016780.D
 Acq On : 16 Jun 2020 15:30
 Operator : JC/SP
 Sample : L2993-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AOC50MW002-200610

Quant Time: Jun 17 04:32:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	229322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	365777	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	374813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	186894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	131664	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
35) Dibromofluoromethane	5.47	113	108841	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
50) Toluene-d8	8.70	98	445602	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
62) 4-Bromofluorobenzene	11.12	95	179723	56.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.92%	
Target Compounds						
16) Acetone	2.42	43	2260	1.932	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061620\
Data File : VX016780.D
Acq On : 16 Jun 2020 15:30
Operator : JC/SP
Sample : L2993-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AOC50MW002-200610

Quant Time: Jun 17 04:32:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

