

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061620\
 Data File : VX016787.D
 Acq On : 16 Jun 2020 18:08
 Operator : JC/SP
 Sample : L3003-01 200X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0570

Quant Time: Jun 17 04:56:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	234270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	369270	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	382852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	196009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	136391	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
35) Dibromofluoromethane	5.46	113	113549	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
50) Toluene-d8	8.70	98	455313	54.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.40%	
62) 4-Bromofluorobenzene	11.12	95	184185	57.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.64%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.15	96	1094	0.445	ug/l #	76
25) 2-Butanone	4.64	43	38093	21.146	ug/l	95
44) Trichloroethene	7.19	130	5121	1.666	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061620\
Data File : VX016787.D
Acq On : 16 Jun 2020 18:08
Operator : JC/SP
Sample : L3003-01 200X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0570

Quant Time: Jun 17 04:56:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

