

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061620\
 Data File : VX016790.D
 Acq On : 16 Jun 2020 19:16
 Operator : JC/SP
 Sample : L2984-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-TB-2020-1

Quant Time: Jun 17 06:29:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	254644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	397889	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	420215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	223962	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	138955	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
35) Dibromofluoromethane	5.46	113	119650	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
50) Toluene-d8	8.70	98	497204	54.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.86%	
62) 4-Bromofluorobenzene	11.12	95	210664	60.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.68%	
Target Compounds						
44) Trichloroethene	7.20	130	1943	0.587	ug/l	82
95) Naphthalene	13.82	128	9089	0.591	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061620\
Data File : VX016790.D
Acq On : 16 Jun 2020 19:16
Operator : JC/SP
Sample : L2984-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-TB-2020-1

Quant Time: Jun 17 06:29:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

