

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061620\
 Data File : VX016797.D
 Acq On : 16 Jun 2020 23:26
 Operator : JC/SP
 Sample : VX0616WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0616WBL02

Quant Time: Jun 17 06:45:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	275787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	432329	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	439222	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226278	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	150156	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
35) Dibromofluoromethane	5.46	113	128863	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
50) Toluene-d8	8.70	98	523343	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
62) 4-Bromofluorobenzene	11.12	95	214627	57.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.10%	
Target Compounds						
44) Trichloroethene	7.20	130	4202	1.168	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061620\
Data File : VX016797.D
Acq On : 16 Jun 2020 23:26
Operator : JC/SP
Sample : VX0616WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0616WBL02

Quant Time: Jun 17 06:45:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

