

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061620\
 Data File : VX016807.D
 Acq On : 17 Jun 2020 03:12
 Operator : JC/SP
 Sample : L2984-05DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0565DL

Quant Time: Jun 17 07:08:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	289468	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	506771	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	471527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	264379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	159140	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
35) Dibromofluoromethane	5.46	113	139253	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
50) Toluene-d8	8.70	98	553934	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
62) 4-Bromofluorobenzene	11.12	95	231155	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
Target Compounds						
16) Acetone	2.42	43	5463	3.699	ug/l	97
25) 2-Butanone	4.64	43	106398	47.801	ug/l	100
27) cis-1,2-Dichloroethene	4.56	96	1318	0.379	ug/l	53
44) Trichloroethene	7.18	130	1909	0.453	ug/l	96
52) Toluene	8.77	92	3245	0.366	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061620\
Data File : VX016807.D
Acq On : 17 Jun 2020 03:12
Operator : JC/SP
Sample : L2984-05DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
S49-0565DL

Quant Time: Jun 17 07:08:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

