

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061620\
 Data File : VX016794.D
 Acq On : 16 Jun 2020 20:47
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 06:38:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	115	0.00
2 T	Dichlorodifluoromethane	0.504	0.539	-6.9	108	0.00
3 P	Chloromethane	0.546	0.513	6.0	111	0.00
4 C	Vinyl Chloride	0.582	0.587	-0.9#	114	0.00
5 T	Bromomethane	0.362	0.307	15.2	106	0.00
6 T	Chloroethane	0.359	0.358	0.3	116	0.00
7 T	Trichlorofluoromethane	0.839	0.825	1.7	108	0.00
8 T	Diethyl Ether	0.343	0.322	6.1	118	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.427	0.446	-4.4	117	0.00
10 T	Methyl Iodide	0.533	0.452	15.2	96	0.00
11 T	Tert butyl alcohol	0.141	0.134	5.0	116	0.00
12 CM	1,1-Dichloroethene	0.489	0.478	2.2#	119	0.00
13 T	Acrolein	0.064	0.057	10.9	117	0.00
14 T	Allyl chloride	0.764	0.738	3.4	117	0.00
15 T	Acrylonitrile	0.287	0.279	2.8	117	0.00
16 T	Acetone	0.255	0.244	4.3	122	0.00
17 T	Carbon Disulfide	1.400	1.313	6.2	113	0.00
18 T	Methyl Acetate	0.681	0.697	-2.3	120	0.00
19 T	Methyl tert-butyl Ether	1.666	1.583	5.0	114	0.00
20 T	Methylene Chloride	0.550	0.512	6.9	116	0.00
21 T	trans-1,2-Dichloroethene	0.525	0.518	1.3	119	0.00
22 T	Diisopropyl ether	1.436	1.395	2.9	115	0.00
23 T	Vinyl Acetate	1.241	1.217	1.9	117	0.00
24 P	1,1-Dichloroethane	0.898	0.857	4.6	114	0.00
25 T	2-Butanone	0.384	0.379	1.3	118	0.00
26 T	2,2-Dichloropropane	0.831	0.710	14.6	101	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.570	5.0	117	0.00
28 T	Bromochloromethane	0.383	0.351	8.4	115	0.00
29 T	Tetrahydrofuran	0.249	0.246	1.2	119	0.00
30 C	Chloroform	0.950	0.883	7.1#	112	0.00
31 T	Cyclohexane	0.671	0.725	-8.0	114	0.00
32 T	1,1,1-Trichloroethane	0.862	0.821	4.8	110	0.00
33 S	1,2-Dichloroethane-d4	0.593	0.520	12.3	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	118	0.00
35 S	Dibromofluoromethane	0.308	0.285	7.5	112	0.00
36 T	1,1-Dichloropropene	0.442	0.434	1.8	113	0.00
37 T	Ethyl Acetate	0.484	0.477	1.4	120	0.00
38 T	Carbon Tetrachloride	0.502	0.486	3.2	110	0.00
39 T	Methylcyclohexane	0.466	0.497	-6.7	115	0.00
40 TM	Benzene	1.345	1.313	2.4	116	0.00
41 T	Methacrylonitrile	0.258	0.254	1.6	119	0.00
42 TM	1,2-Dichloroethane	0.491	0.453	7.7	111	0.00
43 T	Isopropyl Acetate	0.758	0.755	0.4	119	0.00
44 TM	Trichloroethene	0.416	0.511	-22.8	145	0.00
45 C	1,2-Dichloropropane	0.330	0.326	1.2#	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX061620\
 Data File : VX016794.D
 Acq On : 16 Jun 2020 20:47
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 06:38:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.240	0.226	5.8	114	0.00
47 T	Bromodichloromethane	0.501	0.473	5.6	112	0.00
48 T	Methyl methacrylate	0.359	0.364	-1.4	119	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	114	0.00
50 S	Toluene-d8	1.137	1.103	3.0	114	0.00
51 T	4-Methyl-2-Pentanone	0.467	0.476	-1.9	118	0.00
52 CM	Toluene	0.874	0.858	1.8#	116	0.00
53 T	t-1,3-Dichloropropene	0.566	0.543	4.1	113	0.00
54 T	cis-1,3-Dichloropropene	0.590	0.569	3.6	115	0.00
55 T	1,1,2-Trichloroethane	0.351	0.346	1.4	117	0.00
56 T	Ethyl methacrylate	0.526	0.547	-4.0	118	0.00
57 T	1,3-Dichloropropane	0.591	0.570	3.6	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.271	-4.6	117	0.00
59 T	2-Hexanone	0.361	0.369	-2.2	117	0.00
60 T	Dibromochloromethane	0.415	0.413	0.5	115	0.00
61 T	1,2-Dibromoethane	0.383	0.376	1.8	115	0.00
62 S	4-Bromofluorobenzene	0.435	0.417	4.1	114	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	116	0.00
64 T	Tetrachloroethene	0.480	0.473	1.5	115	0.00
65 PM	Chlorobenzene	1.015	1.001	1.4	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.405	0.395	2.5	113	0.00
67 C	Ethyl Benzene	1.715	1.737	-1.3#	115	0.00
68 T	m/p-Xylenes	0.650	0.667	-2.6	114	0.00
69 T	o-Xylene	0.626	0.633	-1.1	113	0.00
70 T	Styrene	1.083	1.131	-4.4	115	0.00
71 P	Bromoform	0.345	0.359	-4.1	119	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
73 T	Isopropylbenzene	3.281	3.277	0.1	114	0.00
74 T	N-amyl acetate	1.275	1.320	-3.5	118	0.00
75 P	1,1,2,2-Tetrachloroethane	0.965	0.922	4.5	117	0.00
76 T	1,2,3-Trichloropropane	1.051	1.006	4.3	117	0.00
77 T	Bromobenzene	0.926	0.887	4.2	116	0.00
78 T	n-propylbenzene	3.563	3.611	-1.3	114	0.00
79 T	2-Chlorotoluene	2.253	2.196	2.5	115	0.00
80 T	1,3,5-Trimethylbenzene	2.750	2.759	-0.3	114	0.00
81 T	trans-1,4-Dichloro-2-butene	0.418	0.390	6.7	112	0.00
82 T	4-Chlorotoluene	2.637	2.585	2.0	116	0.00
83 T	tert-Butylbenzene	2.517	2.520	-0.1	114	0.00
84 T	1,2,4-Trimethylbenzene	2.782	2.834	-1.9	116	0.00
85 T	sec-Butylbenzene	2.944	3.031	-3.0	115	0.00
86 T	p-Isopropyltoluene	2.824	2.898	-2.6	115	0.00
87 T	1,3-Dichlorobenzene	1.566	1.534	2.0	116	0.00
88 T	1,4-Dichlorobenzene	1.624	1.551	4.5	116	0.00
89 T	n-Butylbenzene	2.332	2.406	-3.2	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061620\
Data File : VX016794.D
Acq On : 16 Jun 2020 20:47
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 17 06:38:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.521	0.498	4.4	113	0.00
91 T	1,2-Dichlorobenzene	1.566	1.476	5.7	115	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.257	4.8	115	0.00
93 T	1,2,4-Trichlorobenzene	1.097	1.056	3.7	119	0.00
94 T	Hexachlorobutadiene	0.497	0.457	8.0	112	0.00
95 T	Naphthalene	3.436	3.519	-2.4	118	0.00
96 T	1,2,3-Trichlorobenzene	1.079	1.037	3.9	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6