

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
 Data File : VX022717.D
 Acq On : 16 Jun 2021 13:29
 Operator : JC/MD
 Sample : M2678-11RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-AMND-COMP-D(CL-01A)RE

Quant Time: Jun 17 02:21:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45192	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	107101	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	43468	53.482	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.960%			
35) Dibromofluoromethane	5.391	113	22793	33.356	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 66.720%#			
50) Toluene-d8	8.653	98	138859	51.549	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.100%			
62) 4-Bromofluorobenzene	11.086	95	54950	53.938	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.880%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	10938	30.055	ug/l	99
20) Methylene Chloride	2.794	84	5695	7.870	ug/l	88
25) 2-Butanone	4.574	43	1992	3.375	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
Data File : VX022717.D
Acq On : 16 Jun 2021 13:29
Operator : JC/MD
Sample : M2678-11RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-D(CL-01A)RE

Quant Time: Jun 17 02:21:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

