

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
 Data File : VX022719.D
 Acq On : 16 Jun 2021 14:17
 Operator : JC/MD
 Sample : M2678-13RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-AMND-COMP-G(CL-07)RE

Quant Time: Jun 17 02:21:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	44768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	109148	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46871	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	44630	55.431	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.860%
35) Dibromofluoromethane	5.397	113	20029	28.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	57.520%#
50) Toluene-d8	8.653	98	139832	50.937	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.880%
62) 4-Bromofluorobenzene	11.085	95	57045	54.944	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.880%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6229	17.278	ug/l	89
20) Methylene Chloride	2.794	84	9917	14.046	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
Data File : VX022719.D
Acq On : 16 Jun 2021 14:17
Operator : JC/MD
Sample : M2678-13RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE

Quant Time: Jun 17 02:21:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

