

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
 Data File : VX022721.D
 Acq On : 16 Jun 2021 15:04
 Operator : JC/MD
 Sample : M2678-15RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-AMND-COMPOSITE-HRE

Quant Time: Jun 17 02:22:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	42032	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	100033	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	41018	54.261	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	= 108.520%		
35) Dibromofluoromethane	5.397	113	18057	28.292	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	= 56.580%#		
50) Toluene-d8	8.653	98	133360	53.006	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	= 106.020%		
62) 4-Bromofluorobenzene	11.085	95	52454	55.126	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	= 110.260%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5403	15.962	ug/l	92
20) Methylene Chloride	2.794	84	5136	7.623	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
Data File : VX022721.D
Acq On : 16 Jun 2021 15:04
Operator : JC/MD
Sample : M2678-15RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMPOSITE-HRE

Quant Time: Jun 17 02:22:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

