

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
 Data File : VX022722.D
 Acq On : 16 Jun 2021 15:27
 Operator : JC/MD
 Sample : M2709-16RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-AMND-COMP-N(CL-22)RE

Quant Time: Jun 17 02:22:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	43805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	106135	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	43035	54.625	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.260%			
35) Dibromofluoromethane	5.397	113	23994	35.433	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 70.860%#			
50) Toluene-d8	8.653	98	137954	51.679	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.360%			
62) 4-Bromofluorobenzene	11.085	95	56690	56.152	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 112.300%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	6828	19.356	ug/l	99
20) Methylene Chloride	2.794	84	3037	4.204	ug/l	93
25) 2-Butanone	4.574	43	1483	2.592	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
Data File : VX022722.D
Acq On : 16 Jun 2021 15:27
Operator : JC/MD
Sample : M2709-16RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-N(CL-22)RE

Quant Time: Jun 17 02:22:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

