

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
 Data File : VX022719.D
 Acq On : 16 Jun 2021 14:17
 Operator : JC/MD
 Sample : M2678-13RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-AMND-COMP-G(CL-07)RE

Quant Time: Jun 17 02:21:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.562	168	44768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	109148	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46871	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	44630	55.431	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.860%
35) Dibromofluoromethane	5.397	113	20029	28.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	57.520%#
50) Toluene-d8	8.653	98	139832	50.937	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.880%
62) 4-Bromofluorobenzene	11.085	95	57045	54.944	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.880%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6229	17.278	ug/l	89
20) Methylene Chloride	2.794	84	9917	14.046	ug/l #	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
Data File : VX022719.D
Acq On : 16 Jun 2021 14:17
Operator : JC/MD
Sample : M2678-13RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE

Quant Time: Jun 17 02:21:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17




Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE


Tgt Ion:168 Resp: 44768
Ion Ratio Lower Upper
168 100
99 67.2 45.8 68.6



#16
Acetone
Concen: 17.278 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Tgt Ion: 43 Resp: 6229
Ion Ratio Lower Upper
43 100
58 28.7 27.8 41.8









#20
Methylene Chloride
Concen: 14.046 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE

Tgt Ion: 84 Resp: 9917




Ion	Ratio	Lower	Upper
84	100		
49	141.0	88.6	132.8#
51	37.6	27.0	40.6
86	65.5	50.3	75.5




#33
1,2-Dichloroethane-d4
Concen: 55.431 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Tgt Ion: 65 Resp: 44630

Ion	Ratio	Lower	Upper
65	100		
67	52.0	0.0	105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17




Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE


Tgt Ion:114 Resp: 109148
Ion Ratio Lower Upper
114 100
63 22.0 0.0 40.8
88 17.4 0.0 32.8



#35
Dibromofluoromethane
Concen: 28.761 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Tgt Ion:113 Resp: 20029
Ion Ratio Lower Upper
113 100
111 102.7 81.9 122.9
192 16.4 16.7 25.1#









#50
Toluene-d8
Concen: 50.937 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17




Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE


Tgt Ion: 98 Resp: 139832
Ion Ratio Lower Upper
98 100
100 67.6 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 54.944 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Tgt Ion: 95 Resp: 57045
Ion Ratio Lower Upper
95 100
174 62.1 0.0 154.6
176 59.7 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMND-COMP-G(CL-07)RE

Tgt Ion:117 Resp: 113425
Ion Ratio Lower Upper
117 100
82 63.4 47.4 71.2
119 30.7 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX022719.D
Acq: 16 Jun 2021 14:17

Tgt Ion:152 Resp: 46871
Ion Ratio Lower Upper
152 100
115 68.9 41.1 123.3
150 155.7 0.0 349.0

