



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061621\
 Data File : VX022710.D
 Acq On : 16 Jun 2021 09:08
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Title : SW846 8260
 Last Update : Thu Jun 10 12:20:00 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	4600	PASS
75	95	30	60	49.9	11399	PASS
95	95	100	100	100.0	22842	PASS
96	95	5	9	6.5	1475	PASS
173	174	0.00	2	1.1	152	PASS
174	95	50	100	62.9	14360	PASS
175	174	5	9	7.0	1010	PASS
176	174	95	101	101.0	14502	PASS
177	176	5	9	6.5	944	PASS