

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061623\
 Data File : VX036184.D
 Acq On : 16 Jun 2023 14:47
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 19 01:40:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	124342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210438	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99641	43.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.700%
35) Dibromofluoromethane	5.385	113	84267	59.051	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.100%
50) Toluene-d8	8.647	98	207920	39.832	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	79.660%#
62) 4-Bromofluorobenzene	11.079	95	132425	62.848	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	125.700%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061623\
Data File : VX036184.D
Acq On : 16 Jun 2023 14:47
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 19 01:40:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

