

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061623\
 Data File : VX036195.D
 Acq On : 16 Jun 2023 20:45
 Operator : JC/MD
 Sample : 03301-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1619

Quant Time: Jun 19 08:01:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	168299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	307775	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	276374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	160309	52.123	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.240%			
35) Dibromofluoromethane	5.385	113	103264	49.478	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.960%			
50) Toluene-d8	8.647	98	359900	47.142	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 94.280%			
62) 4-Bromofluorobenzene	11.079	95	137579	44.644	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.280%			
Target Compounds						
16) Acetone	2.386	43	20837	16.307	ug/l	91
20) Methylene Chloride	2.794	84	1179	0.509	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061623\
Data File : VX036195.D
Acq On : 16 Jun 2023 20:45
Operator : JC/MD
Sample : 03301-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1619

Quant Time: Jun 19 08:01:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

