

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061719\
 Data File : VX010322.D
 Acq On : 17 Jun 2019 14:22
 Operator : JC/SP
 Sample : K3154-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-02-061419-E

Quant Time: Jun 18 02:15:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	326096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	490838	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	454679	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	215465	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	144896	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
35) Dibromofluoromethane	5.50	113	147947	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	8.71	98	579649	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	11.13	95	200051	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
Target Compounds						
16) Acetone	2.44	43	17115	11.432	ug/l #	88
20) Methylene Chloride	2.85	84	4277	1.301	ug/l #	83
43) Isopropyl Acetate	6.45	43	11980	1.720	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061719\
Data File : VX010322.D
Acq On : 17 Jun 2019 14:22
Operator : JC/SP
Sample : K3154-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-02-061419-E

Quant Time: Jun 18 02:15:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

