

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061719\
 Data File : VX010324.D
 Acq On : 17 Jun 2019 15:12
 Operator : JC/SP
 Sample : K3121-01MEDL 10X
 Misc : 5.00/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WASTEMEDL

Quant Time: Jun 17 15:42:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	313180	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	469661	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	436647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	215897	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	140731	46.04	ug/l	0.00
Spiked Amount			Recovery	=	92.08%	
35) Dibromofluoromethane	5.49	113	143303	49.47	ug/l	0.00
Spiked Amount			Recovery	=	98.94%	
50) Toluene-d8	8.71	98	561259	51.69	ug/l	0.00
Spiked Amount			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.13	95	201949	49.80	ug/l	0.00
Spiked Amount			Recovery	=	99.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	7.46	83	25461	4.880	ug/l #	81
52) Toluene	8.78	92	23143	2.971	ug/l	97
67) Ethyl Benzene	10.25	91	86247	5.622	ug/l	96
68) m/p-Xylenes	10.35	106	185287	31.210	ug/l	89
69) o-Xylene	10.69	106	109198	18.504	ug/l	90
73) Isopropylbenzene	11.02	105	74514	4.944	ug/l	95
78) n-propylbenzene	11.36	91	220339	13.132	ug/l	95
80) 1,3,5-Trimethylbenzene	11.51	105	282113	22.224	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	972603	75.855	ug/l	95
85) sec-Butylbenzene	11.94	105	180349	12.049	ug/l	85
86) p-Isopropyltoluene	12.06	119	94637	6.916	ug/l	96
89) n-Butylbenzene	12.38	91	311954	26.484	ug/l #	41

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061719\
Data File : VX010324.D
Acq On : 17 Jun 2019 15:12
Operator : JC/SP
Sample : K3121-01MEDL 10X
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WASTEMEDL

Quant Time: Jun 17 15:42:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

