

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061719\
 Data File : VX010327.D
 Acq On : 17 Jun 2019 16:22
 Operator : JC/SP
 Sample : IBLK1
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK1

Quant Time: Jun 18 02:18:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	363195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	545106	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	486855	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	239066	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	172224	48.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.16%	
35) Dibromofluoromethane	5.50	113	165920	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	
50) Toluene-d8	8.71	98	636299	50.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.98%	
62) 4-Bromofluorobenzene	11.13	95	223839	47.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061719\
Data File : VX010327.D
Acq On : 17 Jun 2019 16:22
Operator : JC/SP
Sample : IBLK1
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK1

Quant Time: Jun 18 02:18:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

