

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061719\
 Data File : VX010328.D
 Acq On : 17 Jun 2019 16:45
 Operator : JC/SP
 Sample : IBLK2
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK2

Quant Time: Jun 18 02:19:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	352873	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	528418	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	475216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	225461	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	161714	46.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.90%	
35) Dibromofluoromethane	5.50	113	160322	49.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.38%	
50) Toluene-d8	8.71	98	620650	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.13	95	215778	47.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061719\
Data File : VX010328.D
Acq On : 17 Jun 2019 16:45
Operator : JC/SP
Sample : IBLK2
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK2

Quant Time: Jun 18 02:19:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

