

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016814.D
 Acq On : 17 Jun 2020 14:15
 Operator : JC/SP
 Sample : VX0617WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0617WBL01

Quant Time: Jun 18 01:10:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	313773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	498594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	253997	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	173361	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
35) Dibromofluoromethane	5.46	113	150735	48.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.40%	
50) Toluene-d8	8.70	98	608647	53.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.66%	
62) 4-Bromofluorobenzene	11.12	95	247175	56.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016814.D
Acq On : 17 Jun 2020 14:15
Operator : JC/SP
Sample : VX0617WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0617WBL01

Quant Time: Jun 18 01:10:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

