

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016816.D
 Acq On : 17 Jun 2020 15:07
 Operator : JC/SP
 Sample : L2993-04DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW014-200610DL

Quant Time: Jun 18 01:15:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	292504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	466490	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	535194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230461	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	161028	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
35) Dibromofluoromethane	5.47	113	139601	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
50) Toluene-d8	8.70	98	552755	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
62) 4-Bromofluorobenzene	11.12	95	232246	57.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.42%	
Target Compounds						
19) Methyl tert-butyl Ether	3.16	73	8233	0.845	ug/l	96
21) trans-1,2-Dichloroethene	3.14	96	31988	10.412	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	16285	4.638	ug/l	90
42) 1,2-Dichloroethane	6.16	62	24038	5.247	ug/l	98
44) Trichloroethene	7.19	130	268485	69.152	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016816.D
Acq On : 17 Jun 2020 15:07
Operator : JC/SP
Sample : L2993-04DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW014-200610DL

Quant Time: Jun 18 01:15:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

