

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016817.D
 Acq On : 17 Jun 2020 15:29
 Operator : JC/SP
 Sample : L2993-01RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW039-200610RE

Quant Time: Jun 18 01:18:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	283831	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	495976	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	455444	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	235882	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	158543	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	5.46	113	134978	44.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.22%	
50) Toluene-d8	8.70	98	545376	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	11.12	95	265516	61.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.04%	
Target Compounds						
16) Acetone	2.41	43	3858	2.664	ug/l	99
44) Trichloroethene	7.18	130	1848	0.448	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016817.D
Acq On : 17 Jun 2020 15:29
Operator : JC/SP
Sample : L2993-01RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW039-200610RE

Quant Time: Jun 18 01:18:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

