

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016823.D
 Acq On : 17 Jun 2020 17:43
 Operator : JC/SP
 Sample : L3042-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-200616

Quant Time: Jun 18 01:35:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	254549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	403724	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	411037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	209871	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	142121	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
35) Dibromofluoromethane	5.46	113	119940	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
50) Toluene-d8	8.70	98	563136	61.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.62%	
62) 4-Bromofluorobenzene	11.12	95	201868	57.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.92%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	7386	2.417	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016823.D
Acq On : 17 Jun 2020 17:43
Operator : JC/SP
Sample : L3042-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-200616

Quant Time: Jun 18 01:35:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

