

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016824.D
 Acq On : 17 Jun 2020 18:06
 Operator : JC/SP
 Sample : L3032-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-TB11-200615

Quant Time: Jun 18 01:38:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	255052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	460120	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	410588	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	236967	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	165672	57.74	ug/l	0.00
Spiked Amount			Recovery	=	115.48%	
35) Dibromofluoromethane	5.47	113	121589	42.83	ug/l	0.00
Spiked Amount			Recovery	=	85.66%	
50) Toluene-d8	8.70	98	489158	46.73	ug/l	0.00
Spiked Amount			Recovery	=	93.46%	
62) 4-Bromofluorobenzene	11.12	95	225743	56.38	ug/l	0.00
Spiked Amount			Recovery	=	112.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016824.D
Acq On : 17 Jun 2020 18:06
Operator : JC/SP
Sample : L3032-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
Z1-TB11-200615

Quant Time: Jun 18 01:38:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

