

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016832.D
 Acq On : 17 Jun 2020 21:06
 Operator : JC/SP
 Sample : L2984-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0564

Quant Time: Jun 18 02:06:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	247556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	444196	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	449099	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	138332	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
35) Dibromofluoromethane	5.47	113	116758	42.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.20%	
50) Toluene-d8	8.70	98	462971	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
62) 4-Bromofluorobenzene	11.12	95	206048	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	26732	9.271	ug/l	100
16) Acetone	2.42	43	40337	31.935	ug/l	100
21) trans-1,2-Dichloroethene	3.14	96	1917	0.737	ug/l #	59
25) 2-Butanone	4.63	43	236431	124.203	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	10297	3.465	ug/l	89
29) Tetrahydrofuran	5.09	42	3506	2.843	ug/l #	71
44) Trichloroethene	7.20	130	1690	0.457	ug/l	99
52) Toluene	8.76	92	4550	0.586	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016832.D
Acq On : 17 Jun 2020 21:06
Operator : JC/SP
Sample : L2984-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0564

Quant Time: Jun 18 02:06:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

