

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061720\
 Data File : VX016833.D
 Acq On : 17 Jun 2020 21:29
 Operator : JC/SP
 Sample : L2984-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0566

Quant Time: Jun 18 02:08:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 16 02:02:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	242989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	449797	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	397848	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	216110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	138402	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
35) Dibromofluoromethane	5.47	113	118759	42.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.58%	
50) Toluene-d8	8.70	98	477493	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
62) 4-Bromofluorobenzene	11.12	95	230887	58.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.98%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	1515	0.535	ug/l	92
16) Acetone	2.42	43	50826	40.996	ug/l	100
17) Carbon Disulfide	2.55	76	3646	0.536	ug/l #	79
21) trans-1,2-Dichloroethene	3.15	96	986	0.386	ug/l #	74
25) 2-Butanone	4.64	43	156070	83.528	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	3199	1.097	ug/l	85
44) Trichloroethene	7.20	130	7187	1.920	ug/l	95
52) Toluene	8.76	92	15877	2.020	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061720\
Data File : VX016833.D
Acq On : 17 Jun 2020 21:29
Operator : JC/SP
Sample : L2984-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0566

Quant Time: Jun 18 02:08:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061520W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 16 02:02:35 2020
Response via : Initial Calibration

