

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022726.D
 Acq On : 17 Jun 2021 09:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 18 01:14:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	72033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137968	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57384	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59963	46.29	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.58%
35) Dibromofluoromethane	5.385	113	43421	49.33	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.66%
50) Toluene-d8	8.653	98	172502	49.71	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.42%
62) 4-Bromofluorobenzene	11.085	95	65585	49.97	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.94%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.173	85	42095	49.80	ug/l 97
3) Chloromethane	1.294	50	44086	43.16	ug/l 99
4) Vinyl Chloride	1.374	62	45192	45.03	ug/l 98
5) Bromomethane	1.599	94	27914	51.13	ug/l 93
6) Chloroethane	1.679	64	27917	45.94	ug/l 96
7) Trichlorofluoromethane	1.886	101	69922	48.38	ug/l 99
8) Diethyl Ether	2.136	74	28331	48.52	ug/l 90
9) 1,1,2-Trichlorotrifluo...	2.325	101	44823	50.84	ug/l 90
10) Methyl Iodide	2.453	142	41263	41.74	ug/l 94
11) Tert butyl alcohol	2.977	59	60382	201.30	ug/l 98
12) 1,1-Dichloroethene	2.319	96	42749	48.88	ug/l 96
13) Acrolein	2.239	56	34454	274.09	ug/l 100
14) Allyl chloride	2.666	41	81981	48.49	ug/l 95
15) Acrylonitrile	3.069	53	139433	236.68	ug/l 99
16) Acetone	2.386	43	118664	204.56	ug/l 93
17) Carbon Disulfide	2.514	76	91713	46.60	ug/l 100
18) Methyl Acetate	2.709	43	78160	44.76	ug/l 93
19) Methyl tert-butyl Ether	3.123	73	150012	49.44	ug/l 99
20) Methylene Chloride	2.794	84	52674	47.01	ug/l 92
21) trans-1,2-Dichloroethene	3.093	96	47177	48.68	ug/l 97
22) Diisopropyl ether	3.764	45	163843	48.97	ug/l 92
23) Vinyl Acetate	3.727	43	755688	245.07	ug/l 95
24) 1,1-Dichloroethane	3.611	63	93634	48.22	ug/l 98
25) 2-Butanone	4.562	43	215786	229.37	ug/l 95
26) 2,2-Dichloropropane	4.477	77	77557	52.38	ug/l 98
27) cis-1,2-Dichloroethene	4.489	96	55888	47.33	ug/l 93
28) Bromochloromethane	4.904	49	45613	48.58	ug/l # 79
29) Tetrahydrofuran	5.007	42	127251	232.30	ug/l 88
30) Chloroform	5.099	83	95410	49.30	ug/l 98
31) Cyclohexane	5.471	56	72870	46.63	ug/l 96
32) 1,1,1-Trichloroethane	5.385	97	78386	47.75	ug/l 97
36) 1,1-Dichloropropene	5.696	75	68224	49.63	ug/l 96
37) Ethyl Acetate	4.721	43	83683	47.90	ug/l # 94
38) Carbon Tetrachloride	5.678	117	64506	54.86	ug/l 95
39) Methylcyclohexane	7.385	83	75664	52.76	ug/l 98
40) Benzene	6.038	78	208828	51.06	ug/l 100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022726.D
 Acq On : 17 Jun 2021 09:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 18 01:14:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	46041	49.17	ug/l	94
42) 1,2-Dichloroethane	6.092	62	78166	51.96	ug/l	97
43) Isopropyl Acetate	6.342	43	140433	50.89	ug/l	93
44) Trichloroethene	7.123	130	50197	52.28	ug/l	95
45) 1,2-Dichloropropane	7.434	63	56785	51.46	ug/l	98
46) Dibromomethane	7.580	93	36239	48.80	ug/l #	86
47) Bromodichloromethane	7.824	83	71385	49.46	ug/l	98
48) Methyl methacrylate	7.696	41	66840	50.46	ug/l	89
49) 1,4-Dioxane	7.659	88	24073	954.84	ug/l	94
51) 4-Methyl-2-Pentanone	8.574	43	449596	256.69	ug/l	90
52) Toluene	8.720	92	130352	51.21	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	81887	47.86	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	88006	47.98	ug/l #	90
55) 1,1,2-Trichloroethane	9.153	97	53307	51.18	ug/l	97
56) Ethyl methacrylate	9.116	69	87779	46.75	ug/l #	82
57) 1,3-Dichloropropane	9.311	76	95599	52.78	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	235741	251.07	ug/l	95
59) 2-Hexanone	9.433	43	362806	271.11	ug/l	88
60) Dibromochloromethane	9.525	129	49686	48.66	ug/l	98
61) 1,2-Dibromoethane	9.610	107	54566	52.49	ug/l	99
64) Tetrachloroethene	9.275	164	44699	54.41	ug/l	90
65) Chlorobenzene	10.079	112	138109	51.01	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	47281	48.31	ug/l	96
67) Ethyl Benzene	10.195	91	258987	51.96	ug/l	99
68) m/p-Xylenes	10.305	106	189157	104.78	ug/l	96
69) o-Xylene	10.646	106	92930	53.13	ug/l	97
70) Styrene	10.659	104	159548	49.41	ug/l	97
71) Bromoform	10.805	173	29772	46.67	ug/l #	99
73) Isopropylbenzene	10.963	105	252605	53.61	ug/l	99
74) N-amyl acetate	10.848	43	128514	50.72	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	85281	50.89	ug/l	97
76) 1,2,3-Trichloropropane	11.244	75	101087	66.44	ug/l	87
77) Bromobenzene	11.201	156	52035	51.70	ug/l	84
78) n-propylbenzene	11.305	91	306559	53.87	ug/l	97
79) 2-Chlorotoluene	11.366	91	177507	51.54	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	209474	53.37	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	26814	46.14	ug/l	88
82) 4-Chlorotoluene	11.457	91	206922	52.51	ug/l	96
83) tert-Butylbenzene	11.719	119	192299	52.17	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	212224	53.31	ug/l	97
85) sec-Butylbenzene	11.896	105	260802	54.17	ug/l	97
86) p-Isopropyltoluene	12.012	119	213033	53.60	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	100852	52.79	ug/l	96
88) 1,4-Dichlorobenzene	12.043	146	99112	50.40	ug/l	96
89) n-Butylbenzene	12.335	91	211936	55.69	ug/l	97
90) Hexachloroethane	12.542	117	32340	49.03	ug/l	72
91) 1,2-Dichlorobenzene	12.335	146	96494	51.83	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	17710	45.21	ug/l	77
93) 1,2,4-Trichlorobenzene	13.591	180	60245	53.94	ug/l	95
94) Hexachlorobutadiene	13.725	225	22578	56.37	ug/l	97
95) Naphthalene	13.780	128	233529	47.94	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	58178	53.43	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022726.D
Acq On : 17 Jun 2021 09:20
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Jun 18 01:14:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022726.D
 Acq On : 17 Jun 2021 09:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050


Quant Time: Jun 18 01:14:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050







Tgt Ion:168 Resp: 72033
Ion Ratio Lower Upper
168 100
99 66.1 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 49.80 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.007 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 85 Resp: 42095
Ion Ratio Lower Upper
85 100
87 31.3 16.4 49.2









#3
Chloromethane
Concen: 43.16 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 44086
Ion Ratio Lower Upper
50 100
52 32.5 26.4 39.6



#4
Vinyl Chloride
Concen: 45.03 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 62 Resp: 45192
Ion Ratio Lower Upper
62 100
64 31.3 25.7 38.5









#5
Bromomethane
Concen: 51.13 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 27914
Ion Ratio Lower Upper
94 100
96 88.5 76.4 114.6



#6
Chloroethane
Concen: 45.94 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 64 Resp: 27917
Ion Ratio Lower Upper
64 100
66 34.0 25.4 38.0









#7
Trichlorofluoromethane
Concen: 48.38 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 69922
Ion Ratio Lower Upper
101 100
103 65.7 53.2 79.8



#8
Diethyl Ether
Concen: 48.52 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 74 Resp: 28331
Ion Ratio Lower Upper
74 100
45 97.9 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.84 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 44823
Ion Ratio Lower Upper
101 100
85 47.0 34.9 52.3
151 66.0 61.4 92.2



#10
Methyl Iodide
Concen: 41.74 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:142 Resp: 41263
Ion Ratio Lower Upper
142 100
127 46.6 33.3 49.9
141 14.1 11.4 17.0









#11

Tert butyl alcohol
Concen: 201.30 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 60382
Ion Ratio Lower Upper
59 100
57 10.4 7.8 11.8




#12

1,1-Dichloroethene
Concen: 48.88 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 96 Resp: 42749
Ion Ratio Lower Upper
96 100
61 160.9 124.6 187.0
98 58.9 50.6 76.0









#13
Acrolein
Concen: 274.09 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 34454
Ion Ratio Lower Upper
56 100
55 70.1 56.4 84.6



#14
Allyl chloride
Concen: 48.49 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 41 Resp: 81981
Ion Ratio Lower Upper
41 100
39 65.8 51.6 77.4
76 31.8 30.5 45.7









#15
Acrylonitrile
Concen: 236.68 ug/l
RT: 3.069 min Scan# 325
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	53	Resp:	139433
Ion	Ratio	Lower	Upper
53	100		
52	82.1	65.3	97.9
51	36.0	28.2	42.2



#16
Acetone
Concen: 204.56 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:	43	Resp:	118664
Ion	Ratio	Lower	Upper
43	100		
58	30.9	27.8	41.8









#17
Carbon Disulfide
Concen: 46.60 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 91713
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8



#18
Methyl Acetate
Concen: 44.76 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 43 Resp: 78160
Ion Ratio Lower Upper
43 100
74 23.8 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 49.44 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 150012
Ion Ratio Lower Upper
73 100
57 22.1 17.1 25.7



#20
Methylene Chloride
Concen: 47.01 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 84 Resp: 52674
Ion Ratio Lower Upper
84 100
49 125.4 88.6 132.8
51 36.3 27.0 40.6
86 62.0 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 48.68 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 47177
Ion Ratio Lower Upper
96 100
61 137.3 107.7 161.5
98 60.2 50.4 75.6



#22
Diisopropyl ether
Concen: 48.97 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 45 Resp: 163843
Ion Ratio Lower Upper
45 100
43 78.6 58.0 87.0
87 26.6 25.4 38.0
59 10.2 10.2 15.2









#23
 Vinyl Acetate
 Concen: 245.07 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.000 min
 Lab File: VX022726.D
 Acq: 17 Jun 2021 09:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 755688
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 48.22 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX022726.D
 Acq: 17 Jun 2021 09:20

Tgt Ion: 63 Resp: 93634
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.8 11.2
 100 5.3 2.5 7.5





#25

2-Butanone

Concen: 229.37 ug/l

RT: 4.562 min Scan# 570

Delta R.T. -0.006 min

Lab File: VX022726.D





Acq: 17 Jun 2021 09:20

Tgt Ion: 43 Resp: 215786

Ion Ratio Lower Upper

43 100

72 26.7 23.7 35.5



#26

2,2-Dichloropropane

Concen: 52.38 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.012 min

Lab File: VX022726.D




Acq: 17 Jun 2021 09:20


Tgt Ion: 77 Resp: 77557

Ion Ratio Lower Upper

77 100

97 22.3 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 47.33 ug/l

RT: 4.489 min Scan# 558





Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion: 96 Resp: 55888

Ion	Ratio	Lower	Upper
96	100		
61	154.4	0.0	285.6
98	65.2	0.0	129.6



#28

Bromochloromethane

Concen: 48.58 ug/l

RT: 4.904 min Scan# 626




Delta R.T. -0.000 min


Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion: 49 Resp: 45613

Ion	Ratio	Lower	Upper
49	100		
129	1.4	0.0	4.0
130	65.9	68.2	102.2#





#29

Tetrahydrofuran

Concen: 232.30 ug/l

RT: 5.007 min Scan# 643

Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20





Tgt Ion: 42 Resp: 127251

Ion Ratio Lower Upper

42 100

72 43.5 41.7 62.5

71 40.3 38.4 57.6



#30

Chloroform

Concen: 49.30 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX022726.D




Acq: 17 Jun 2021 09:20


Tgt Ion: 83 Resp: 95410

Ion Ratio Lower Upper

83 100

85 67.0 52.0 78.0





#31

Cyclohexane

Concen: 46.63 ug/l

RT: 5.471 min Scan# 719





Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion:	56	Resp:	72870
Ion	Ratio	Lower	Upper
56	100		
69	28.8	25.3	37.9
84	85.9	71.4	107.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



#32

1,1,1-Trichloroethane

Concen: 47.75 ug/l




RT: 5.385 min Scan# 705


Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion:	97	Resp:	78386
Ion	Ratio	Lower	Upper
97	100		
99	63.9	51.5	77.3
61	47.2	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 46.29 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion: 65 Resp: 59963

Ion Ratio Lower Upper

65 100





67 51.1 0.0 105.4

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20




Tgt Ion: 114 Resp: 137968


Ion Ratio Lower Upper

114 100

63 21.0 0.0 40.8

88 17.7 0.0 32.8









#35
Dibromofluoromethane
Concen: 49.33 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 43421
Ion Ratio Lower Upper
113 100
111 104.9 81.9 122.9
192 15.2 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 49.63 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 75 Resp: 68224
Ion Ratio Lower Upper
75 100
110 33.0 18.1 54.4
77 32.6 25.3 37.9





#37

Ethyl Acetate

Concen: 47.90 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20





Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43	Resp: 83683
Ion Ratio	Lower Upper
43 100	
61 13.6	12.8 19.2
70 10.1	10.5 15.7#



#38

Carbon Tetrachloride

Concen: 54.86 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.000 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion: 117	Resp: 64506
Ion Ratio	Lower Upper
117 100	
119 90.1	76.2 114.2
121 27.7	24.8 37.2









#39
Methylcyclohexane
Concen: 52.76 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	83	Resp:	75664
Ion	Ratio	Lower	Upper
83	100		
55	76.6	58.9	88.3
98	47.6	38.1	57.1



#40
Benzene
Concen: 51.06 ug/l
RT: 6.038 min Scan# 812
Delta R.T. -0.012 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:	78	Resp:	208828
Ion	Ratio	Lower	Upper
78	100		
77	23.4	18.9	28.3









#41
Methacrylonitrile
Concen: 49.17 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	46041
Ion	Ratio	Lower	Upper
41	100		
39	52.9	42.5	63.7
67	71.6	65.0	97.6
52	30.8	25.3	37.9



#42
1,2-Dichloroethane
Concen: 51.96 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:	62	Resp:	78166
Ion	Ratio	Lower	Upper
62	100		
98	8.8	0.0	20.2









#43
Isopropyl Acetate
Concen: 50.89 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	43	Resp:	140433
Ion	Ratio	Lower	Upper
43	100		
61	19.4	18.2	27.4
87	12.7	12.5	18.7



#44
Trichloroethene
Concen: 52.28 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:	130	Resp:	50197
Ion	Ratio	Lower	Upper
130	100		
95	103.5	0.0	198.0





#45

1,2-Dichloropropane

Concen: 51.46 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.006 min

Lab File: VX022726.D





Acq: 17 Jun 2021 09:20

Tgt Ion: 63 Resp: 56785

Ion Ratio Lower Upper

63 100

65 31.2 25.7 38.5



#46

Dibromomethane

Concen: 48.80 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20




Tgt Ion: 93 Resp: 36239


Ion Ratio Lower Upper

93 100

95 83.5 66.8 100.2

174 78.4 83.1 124.7#









#47
Bromodichloromethane
Concen: 49.46 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83	Resp: 71385
Ion Ratio	Lower Upper
83 100	
85 62.4	51.4 77.0
127 7.5	6.8 10.2



#48
Methyl methacrylate
Concen: 50.46 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 41	Resp: 66840		
Ion Ratio	Lower	Upper	
41	100		
69	80.7	76.0	114.0
39	60.0	44.6	67.0





#49

1,4-Dioxane

Concen: 954.84 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.006 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion: 88 Resp: 24073

Ion Ratio Lower Upper

88 100

43 34.5 23.2 34.8





58 72.6 55.0 82.4

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050



#50

Toluene-d8

Concen: 49.71 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX022726.D




Acq: 17 Jun 2021 09:20


Tgt Ion: 98 Resp: 172502

Ion Ratio Lower Upper

98 100

100 65.5 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 256.69 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 449596
Ion Ratio Lower Upper
43 100
58 37.9 35.4 53.0



#52
Toluene
Concen: 51.21 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 92 Resp: 130352
Ion Ratio Lower Upper
92 100
91 171.5 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 47.86 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 81887
Ion Ratio Lower Upper
75 100
77 32.4 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 47.98 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 75 Resp: 88006
Ion Ratio Lower Upper
75 100
77 33.6 25.7 38.5
39 53.4 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 51.18 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 97	Resp: 53307
Ion Ratio	Lower Upper
97 100	
83 89.3	69.3 103.9
85 57.3	44.3 66.5
99 60.9	50.3 75.5



#56
Ethyl methacrylate
Concen: 46.75 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 69	Resp: 87779
Ion Ratio	Lower Upper
69 100	
41 70.5	46.8 70.2#
39 42.1	24.7 37.1#









#57

1,3-Dichloropropane
Concen: 52.78 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 95599
Ion Ratio Lower Upper
76 100
78 32.5 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 251.07 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 63 Resp: 235741
Ion Ratio Lower Upper
63 100
106 26.9 23.5 35.3









#59
2-Hexanone
Concen: 271.11 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 362806
Ion Ratio Lower Upper
43 100
58 53.0 30.9 92.8



#60
Dibromochloromethane
Concen: 48.66 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 129 Resp: 49686
Ion Ratio Lower Upper
129 100
127 78.3 38.4 115.2









#61
1,2-Dibromoethane
Concen: 52.49 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 54566
Ion Ratio Lower Upper
107 100
109 93.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 49.97 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 95 Resp: 65585
Ion Ratio Lower Upper
95 100
174 63.8 0.0 154.6
176 60.3 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	117	Resp:	129904
Ion Ratio	Lower	Upper	
117	100		
82	61.4	47.4	71.2
119	32.6	26.6	39.8



#64
Tetrachloroethene
Concen: 54.41 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:	164	Resp:	44699
Ion Ratio	Lower	Upper	
164	100		
166	120.7	103.4	155.2
129	105.0	70.8	106.2
131	97.7	72.5	108.7









#65
Chlorobenzene
Concen: 51.01 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 138109
Ion Ratio Lower Upper
112 100
114 31.8 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 48.31 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:131 Resp: 47281
Ion Ratio Lower Upper
131 100
133 92.8 47.7 143.1
119 68.3 32.1 96.5









#67
Ethyl Benzene
Concen: 51.96 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 258987
Ion Ratio Lower Upper
91 100
106 30.5 24.6 37.0



#68
m/p-Xylenes
Concen: 104.78 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:106 Resp: 189157
Ion Ratio Lower Upper
106 100
91 209.8 162.8 244.2









#69
o-Xylene
Concen: 53.13 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 92930
Ion Ratio Lower Upper
106 100
91 220.1 108.1 324.2



#70
Styrene
Concen: 49.41 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:104 Resp: 159548
Ion Ratio Lower Upper
104 100
78 53.4 40.0 60.0
103 53.8 43.4 65.0





#71

Bromoform

Concen: 46.67 ug/l

RT: 10.805 min Scan# 1594

Delta R.T. -0.000 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20

Tgt Ion:173 Resp: 29772

Ion Ratio Lower Upper

173 100

175 48.8 23.9 71.8





254 0.0 0.1 0.1#

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX022726.D

Acq: 17 Jun 2021 09:20




Tgt Ion:152 Resp: 57384


Ion Ratio Lower Upper

152 100

115 95.3 41.1 123.3

150 175.2 0.0 349.0









#73
Isopropylbenzene
Concen: 53.61 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 252605
Ion Ratio Lower Upper
105 100
120 25.6 13.2 39.6



#74
N-ethyl acetate
Concen: 50.72 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 43 Resp: 128514
Ion Ratio Lower Upper
43 100
70 40.0 39.0 58.6
55 22.8 21.5 32.3
61 23.4 22.1 33.1









#75
 1,1,2,2-Tetrachloroethane
 Concen: 50.89 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX022726.D
 Acq: 17 Jun 2021 09:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 83 Resp: 85281
 Ion Ratio Lower Upper
 83 100
 131 8.7 5.1 15.2
 85 65.6 31.8 95.4



#76
 1,2,3-Trichloropropane
 Concen: 66.44 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX022726.D
 Acq: 17 Jun 2021 09:20

Tgt Ion: 75 Resp: 101087
 Ion Ratio Lower Upper
 75 100
 77 30.7 19.1 57.5









#77
Bromobenzene
Concen: 51.70 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 52035
Ion Ratio Lower Upper
156 100
77 198.5 82.8 248.6
158 96.4 49.1 147.4



#78
n-propylbenzene
Concen: 53.87 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 91 Resp: 306559
Ion Ratio Lower Upper
91 100
120 21.4 11.5 34.5









#79
2-Chlorotoluene
Concen: 51.54 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 177507
Ion Ratio Lower Upper
91 100
126 31.7 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 53.37 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 105 Resp: 209474
Ion Ratio Lower Upper
105 100
120 46.3 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 46.14 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	75	Resp:	26814
Ion	Ratio	Lower	Upper
75	100		
53	106.1	74.5	111.7
89	52.7	37.9	56.9



#82
4-Chlorotoluene
Concen: 52.51 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:	91	Resp:	206922
Ion	Ratio	Lower	Upper
91	100		
126	27.3	14.9	44.5









#83
tert-Butylbenzene
Concen: 52.17 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 192299
Ion Ratio Lower Upper
119 100
91 66.2 27.9 83.7
134 22.6 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 53.31 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:105 Resp: 212224
Ion Ratio Lower Upper
105 100
120 43.2 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 54.17 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022726.D
 Acq: 17 Jun 2021 09:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 260802
 Ion Ratio Lower Upper
 105 100
 134 18.5 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 53.60 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022726.D
 Acq: 17 Jun 2021 09:20

Tgt Ion:119 Resp: 213033
 Ion Ratio Lower Upper
 119 100
 134 24.8 12.9 38.7
 91 26.7 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 52.79 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 100852
Ion Ratio Lower Upper
146 100
111 46.2 20.2 60.5
148 63.3 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 50.40 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:146 Resp: 99112
Ion Ratio Lower Upper
146 100
111 44.7 19.6 58.8
148 63.5 31.6 95.0









#89
n-Butylbenzene
Concen: 55.69 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 211936
Ion Ratio Lower Upper
91 100
92 53.1 27.4 82.0
134 23.3 13.2 39.5



#90
Hexachloroethane
Concen: 49.03 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 117 Resp: 32340
Ion Ratio Lower Upper
117 100
201 55.8 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 51.83 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 96494
Ion Ratio Lower Upper
146 100
111 47.7 20.8 62.2
148 62.6 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 45.21 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion: 75 Resp: 17710
Ion Ratio Lower Upper
75 100
155 70.1 45.3 135.9
157 88.7 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 53.94 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 60245
Ion Ratio Lower Upper
180 100
182 91.2 47.7 143.1
145 34.6 15.0 45.0



#94
Hexachlorobutadiene
Concen: 56.37 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.006 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:225 Resp: 22578
Ion Ratio Lower Upper
225 100
223 58.8 31.1 93.2
227 62.6 31.6 95.0









#95
Naphthalene
Concen: 47.94 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 233529
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 53.43 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022726.D
Acq: 17 Jun 2021 09:20

Tgt Ion:180 Resp: 58178
Ion Ratio Lower Upper
180 100
182 93.4 47.6 142.9
145 37.7 16.1 48.3

